A-A-50893A
16 September 1988
SUPERSEDING
A-A-50893
28 July 1983

COMMERCIAL ITEM DESCRIPTION

LIGHT, DIAGNOSTIC - EXAMINATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a hand held light with an integral magnifier lens, suitable for use in medical examination and diagnosis.

Salient characteristics.

The hand held light shall provide long wave ultraviolet light and day light flourescent illumination with operator observation through an integral magnifier lens.

The light unit shall operate on a nominal 115 volt, 50/60 Hz AC power. The unit shall be provided with a power cord approximately 6 feet long with a hospital grade plug. The unit shall meet applicable portions of UL544.

The weight of the unit shall not exceed 3-1/2 pounds.

The main body of the unit shall contain the following components:

Sockets, ballast, and wiring for four lamps.

Two ultraviolet lamps, 4 watts, 6 inch miniature bipin, Type F4T5BLB.

Two cold white fluorescent lamps, 4 watts, 6 inch miniature bipin, Type F4T5D.

Switch or Switches to provide ON-OFF control and selection of either ultraviolet or white light, flourescent light.

Magnifier Lens.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:48Z
Check the source to verify that this is the current version before use.

The main body of the unit shall consist of a rectangular housing approximately 7 inches long by 5-1/2 inches wide by 2-1/2 inches thick. A handle approximately 6 inches long shall be attached midway on the long side of the main body housing. The magnifying lens shall be centrally mounted in the main body housing with its long plane parallel to the long plane of the housing and at a right angle to the handle bracket. Lamps shall be mounted parallel to the long plane of the magnifier lens and arranged to provide one lamp of each type on each side of the lens.

Lamps shall be in accordance with ANSI Standard C78.2. Ultraviolet light shall be long wave, above 3200 Angstrom units in the ultraviolet spectrum, with peak average of 3500 to 3700 Angstrom units.

The magnifier lens shall provide distortion free viewing of at least 3 power magnification. The lens shall be optically ground and polished glass measuring approximately 4 inches long by 2 inches wide with a nominal 8 inch focal length.

Each unit shall be furnished with two copies of a technical manual containing complete instructions for operation and maintenence, a wiring diagram, illustrations and a listing of component parts. Manufacturer's part numbers shall be given for each component other than common fasteners.

Workmanship. The light shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One diagnostic-examination light, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract of purchase order, the contractor may use his own or nay facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspection set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling. Sampling for tests shall be conducted in accordance with MIL-STD-105. Inspection shall be conducted at Inspection level II with an AQL of 2.5 per 100 units. The unit of product for sampling purposes shall be one diagnostic-examination light.

Tests. Tests shall be conducted to determine compliance with specified requirements. Failure of any test is considered a major defect. Testing shall include but not limited to the following:

The unit shall be connected to a 115 volt 50 Hz power source. Both the daylight and ultraviolet lamps shall be checked for proper operation. Verify the ultraviolet light produces light above 3200 Angstrom units. Repeat the procedure for 60 Hz power source.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6530-00-480-8286	LIGHT, DIAGNOSTIC-EXAMINATION

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the currentversion before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2027

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10173

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50893A	LIGHT, DIAGNOSTIC - EXAMINATION

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.