A-A-50893B
15 June 1990
SUPERSEDING
A-A-50893A
16 September 1988

COMMERCIAL ITEM DESCRIPTION

LIGHT, DIAGNOSTIC-EXAMINATION

(Hand Held)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-L-37216 which is cancelled.

This Commercial Item Description covers a hand held diagnostic-examination light.

Salient Characteristics:

The diagnostic-examination light shall be a hand held light unit which provides long wave ultraviolet light and daylight fluorescent illumination, with operator observation capability through an integral magnifier lens. The weight of the light shall not exceed 3-1/2 pounds.

The light shall operate on a nominal 115 volts, 50/60 Hz. It shall be provided with a 3 conductor power cord with ground lead approximately 6 feet long and terminating in a male, 2-pole, grounded hospital grade plug that complies with American National Standards Institute (ANSI) Standard C73.11. The cord shall be provided with strain relief such that a 35 pound pull on the cord is not transmitted to terminals, splices, or internal wiring when tested in accordance with UL Standard 65, Electric Wired Cabinets. The light shall meet the requirements of the Underwriters' Laboratories, Inc. (UL). The light shall show no signs of dielectric breakdown when tested for dielectric withstand as specified in UL 153.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.

The main body of the light shall consist of a rectangular housing, approximately 7 inches long by 5-1/2 inches wide by 2-1/2 inches thick, containing the following:

a. Sockets, ballast, and wiring for four lamps.

b. Two ultraviolet lamps, 4 watts, 6 inch miniature bipin, type F4T5/BLB.

c. Two daylight white fluorescent lamps, 4 watts, 6 inch miniature bipin, type 54T5D.

d. Switch or switches to provide ON-OFF control and selection of either ultraviolet or daylight fluorescent light.

e. Magnifier lens.

A handle approximately 6 inches long shall be attached midway on the long side of the main body housing. The magnifying lens shall be centrally mounted in the main body housing with its long plane parallel to the long plane of the housing and at a right angle to the handle bracket. The magnifier lens shall provide distortion free viewing of at least 3 power magnification. The lens shall be an optically ground and polished glass measuring approximately 4 inches long by 2 inches wide with a nominal 8 inch focal length.

The daylight lamps shall be in accordance with ANSI Standard C78.2. The ultraviolet lamps shall be long wave, above 3200 Angstrom units in the ultraviolet spectrum, with peak average of 3500 to 3700 Angstrom units. The lamps shall be mounted parallel to the long plane of the magnifier lens and arranged to provide one lamp of each type on each side of the lens.

Two switches shall be provided. Where only one ballast is used, the switches shall be wired in such a manner as to allow only two lamps to be energized at one time. Where two ballasts are provided, one switch shall control one ballast.

The unit shall be supplied with either one or two ballasts. The ballast shall be capable of providing power to two lamps

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the currentversion before use.

at one time. Where one ballast is used, it shall be wired through the switches in such a way that actuating one switch shall energize the two daylight tubes, and actuating the other switch shall energize the two ultraviolet tubes. Where two ballasts are provided, one switch and one ballast shall control the ultraviolet lamps. Starters used in conjunction with the ballast shall not be acceptable. The light shall successfully pass an operational test.

The light reflector shall be fabricated of aluminum sheet or carbon steel sheet. Aluminum shall be anodized aluminum reflector sheet polished to a specular finish. Steel shall be chrome over nickel plated.

Each diagnostic-examination light shall be furnished with two copies of a technical manual containing complete instructions for operation and maintenance, a wiring diagram, illustrations, and a listing of component parts. Manufacturer's part numbers shall be given for each component other than common fasteners.

Workmanship. The diagnostic-examination light shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One diagnostic-examination light, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one diagnostic-examination light.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, with an AQL of 0 (percent defective) and an inspection level of S-4. Sampling for dimensional examination shall be conducted in accordance with Mil-STD-105 with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for both of the above shall be one diagnostic-examination light.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented be Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.

NOTES: 1. The following National Stock Number is covered by this document.

6530-00-480-8280

2. The light shall be suitable for use with NSN 6240-00-319-4549, Lamp, Fluorescent, Ultraviolet, and NSN 6240-00-796-6240, Lamp, Fluorescent, Daylight.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6530-2148

Location: LIGHTCID.13A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.