COMMERCIAL ITEM DESCRIPTION

STERILIZER, DRY HEAT

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-S-36366.

This commercial item description covers a portable, dry heat sterilizer suitable for sterilizing dental and surgical instruments in medical treatment facilities.

Salient characteristics:

The sterilizer shall meet the applicable requirements of UL-544 Standard for Safety for Medical and Dental Equipment. The sterilizer shall be fabricated of stainless steel except for the trays which shall be aluminum. Handles, knobs and other nonmetal parts shall be heat-resistant plastic.

The overall outside dimensions shall not exceed 16-1/2 inches (41.9 cm) wide by 11-1/2 inches (29.2 cm) high by 11 inches (27.9 cm) deep. The dimensions of the sterilizer chamber shall be a minimum of 12 inches (30.5 cm) wide by 5 inches (12.7 cm) high by 7 inches (17.8 cm) deep. The sterilizer shall be supplied with three trays. Two of the trays shall be suitable for holding instruments, and shall have minimum inside dimensions of 9 inches (22.9 cm) wide by 6 inches (15.2 cm) deep by 3/8 inch (1.0 cm) high. The bottom of these two trays shall be perforated. The third tray shall be non-perforated and suitable for holding needles. It shall have minimum inside dimensions of 1-1/4 inches (3.2 cm) wide by 4-5/8 inches (11.7 cm) deep by 3/8 inch (1.0 cm) high. The sterilization chamber opening shall be provided with a snugly fitted door, designed to permit reasy insertion and removal of the instrument trays. The door shall have a suitabrle latch and a gasket to prevent loss of heat. The housing shall be mourned on four rubber or plastic bumpers.

The sterilizer shall be provided with a thermometer, signal light, thermostat and automatic timer. The thermometer shall have a range of at least 300° to 375°F (148.9^{o} to 190.6^{o}C). The adjustable thermostat shall maintain the average temperature of the sterilization chamber to within $\pm$ 10°F ($\pm$ 5.6^{o}C) between 320° and 360°F (160° and 182.2^{o}C). The timer shall have a minimum range of 0 to 120 minutes for automatic operation, and a manual position for continuous operation. The timer shall turn heaters on when it is set to the desired time and off when set time has elapsed. The pilot light shall indicate when the timer is activated. In an ambient temperature of 72°F (22.2^{o}C), the sterilizer shall reach operating temperature of 320°F (160°C) in not more than 20 minutes.

The sterilizer shall operate on 120 volts AC, 60 Hz, at 450 watts. The sterilizer shall be equipped with a three-conductor cord, not less than six feet long. The power cord shall be equipped with a hospital grade plug conforming to ANSI Std. C73.11, and a strain relief in accordance with UL-544. The housing of the sterilizer shall be insulated so that when operating at 360°F (182.2^{o}C), the exterior surface of the sterilizer shall not be hot enough to burn personnel.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the correct version before use.

Each sterilizer shall be furnished with two copies each of a technical manual, which contains complete instructions for operation, maintenance, and a list of component parts identified on an illustration and by manufacturer's part number other than common hardware.

Each sterilizer shall be permanently and legibly marked with the following information:

Contract number
Name or registered trademark of manufacturer
Manufacturer's model number
National Stock Number
Electrical characteristics (volts, amperes, watts)
Serial number

Workmanship. The sterilizer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One sterilizer,as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-962-9965

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PGC 50543

Project No. 6530-1567

Location CID-H1-2-15

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.