A-A-50905
21 January 1985

COMMERCIAL ITEM DESCRIPTION

TUBE, SUCTION, SURGICAL, Sterile

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a plastic surgical suction tube with an attached precision-molded Poole type aspirating end piece.

Salient characteristics:

Shall be a sterile, disposable, clear, transparent suction tube with a universal rigid plastic female connector permanently attached to one end and a Poole-type suction device permanently attached to the opposite end. The female connector shall be capable of fitting standard hospital suction equipment. The suction device shall be approximately 10-1/2 inches long and shall have three-position versatility, governed by a single twist, lock and slide motion. The plastic tubing shall be approximately 72 inches long and shall have a nominal 0.258 inch ID and 0.360 inch OD. Shall comply with Class II requirements for plastics. Shall be tested for sterility in accordance with the USP

Workmanship. The suction tube shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with the requirements specified.

Unit Each (EA). Each suction tube, as specified, constitutes one unit

The suction tubes shall be individually packaged sterile in a peel-open type container capable of maintaining sterility of contents

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheter.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U. S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical

Preservation, packaging, packing labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-068-4898

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

44553

Project No. 6515-2298

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.