A-A-50906
8 May 1985

COMMERCIAL ITEM DESCRIPTION

KNIFE, PLASTER

The General Services Administration has authorized this commercial item description in preference to Interim Federal Specification GG-K-00508C (DSA-DM).

This commercial item description covers a dental knife suitable for use in trimming gypsum casts and in separating dental flasks.

Salient characteristics:

The knife shall consist of a one-piece construction, carbon or alloy steel blade, tapered and ground to a straight cutting edge, forming a knife on the long end and a chisel at the opposite, short end, as shown in Figure 1. The long steel blade (knife) shall be suitable for trimming gypsum casts. The chisel end shall be suitable for separating dental flasks. The handle shall be of a two-piece hardwood and shall be securely pinned through the opposite end with solid, corrosion-resistant, metal pins, and shall be treated to resist absorption or moisture. Overall length shall be 6-5/8 ± 1/4 inches long; blade shall be 2-1/8 ± 1/4 inches long; and the chisel end shall be 1/2 + 1/8 inch long by 11/32 ± 1/16 inch wide. The cross-sectional thickness of the back end of long blade shall be 1/8 ± 1/64 inch. The cross-sectional width of the knife handle, with respect to the long axis of the blade, shall be 3/4 ± 1/8 inch. Rockwell hardness of the blade shall be C-58 to C-62.

Finish. The blade shall be sharpened to form a sharp cutting edge and the chisel end shall be smoothly finished. Wood handles shall have a smooth finish.

Identification marking. The blade of each knife shall be permanently and legibly marked with the name or trademark of the manufacturer or supplier.

Workmanship. The knife shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One knife constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection ?rrequirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more test characteristics. The unit of product for inspection purposes shall be one knife. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities, labeling, and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-00-537-3000

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

Military interests:

Custodians

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY

VA-OSS

Preparing activity:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1621

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.




FIGURE 1

The knife shown is typical in style and design and is not intended to restrict or preclude the purchase of knives which are otherwise in accordance with the requirements of this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.