A-A-50907
13 May 1985

COMMERCIAL ITEM DESCRIPTION

PLASTIC STRIP, DENTAL MATRIX (Polyethylene Terephthalate, Clear, 45 feet)

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification L-P-00525b (DSA-DM).

This commercial item description covers a clear, plastic strip suitable for use as a matrix material during placement of dental restorative materials.

Salient characteristics:

The plastic strip shall be suitable for use as a matrix material during placement of direct resin filling material.

The plastic strip shall be furnished in a continuous roll and shall have the following dimensions: length of strip shall be 45 feet, minimum; width shall be 3/8 + 1/32 inch; thickness shall be 0.0022 + 0.0004 inch; the outside diameter of the roll shall be 1.5 inches, maximum; the inside diameter shall be 5/16 inch, minimum.

ne rolls shall fit and properly function in the dental strip dispenser (NSN 6520-00-523-6050). The strip shall be furnished clean and ready for use.

Material. The plastic strip shall be fabricated of a clear, polyester film (polyethylene terephthalate) sufficiently rigid for use as a matrix material.

Workmanship. The plastic strip shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Roll (RO). One roll containing 45 feet of plastic strip, as specified, constitutes one unit.

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:48Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one plastic strip. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-00-823-8171

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY:

VA - OSS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Project No. 6520-1802

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.