A-A-50910
29 January 1985

COMMERCIAL ITEM DESCRIPTION

HANGER, X-RAY FILM DEVELOPING AND DRYING

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-H-1098.

This commercial item description covers two types of hangers in various sizes used for processing medical and dental X-ray films in developing and drying procedures. Type I comprises medical X-ray film hangers in: Size 1 - 8 by 10 inches, Size 2 - 10 by 12 inches, and Size 3 - 14 by 17 inches. Type II is a dental X-ray film hanger.

Salient characteristics:

All parts of the hangers (except identification tabs for Type II hangers), including clips, clip points, top bar pins, springs, frames, and any other parts, shall be made entirely of corrosion-resisting steel, conforming to the requirements for AISI 302, 303, 304, 304L, 308, or 316 of Fed. Std. No. 66.

Identification tabs for Type II dental X-ray film hangers, shall consist of a suitable material with a white matte finish which will readily accept pen and pencil marks and which will not show any deterioration when subject to developer and fixer solutions.

Each Type I hanger shall consist of a top bar, a rectangular frame, spring clips, and, if required, a crossbar designed and constructed to hold film in a taut, vertical position for processing.

The top bar shall be 16 ± 1/16 inches long of 1/4-inch channel or solid bar stock. If the top bar is a channel, it shall be formed of steel 0.031 inch thick with commercial tolerances. The top bar shall be 0.250 ± 0.015 inch by 0.250 ± 0.015 inch. Each leg of the channel shall form an angle of 90°, $+5^{\circ}$, -0° with the base. The sides of the top bar shall be parallel with the plane of the wire frame, with a maximum offset of 0.005 inch.

For sizes 1 and 2, there shall be a pin 5/8 inch ± 1/8 inch long attached to the under surface of one end of the top bar 3/4 inch ± 1/8 inch from one end and 14-3/8 inches ± 1/8 inch from the opposite side of the frame, outside-to-outside dimensions. The other end of the top bar shall extend 3/4 inch ± 1/8 inch beyond the frame. For hanger sizes 1, 2 and 3, the frame shall be constructed from 3/16 inch diameter rod (commercial tolerances permitted), securely riveted or welded to the top bar, and shall be formed to hold film with its long axis vertical. To assure complete tank submersion of film secured in the hanger, the design for sizes 1 and 2 shall be such that the upper clips fall at least 3 inches below the top bar with a film mounted into position. For size 3 hanger, the two top spring clip assemblies shall be securely spot welded to the under-surface of the top bar. Dimensional requirements shall be as indicated in table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

TABLE I. Frame, dimensional requirements

Size	Overall vertical height (inches)	Inside width of frame (inches)	Outside width of frame (inches)
1	14-1/4 minimum	8 mininum	-----
2	16-7/16 minimum	10 mininum	-----
3	19-7/16 ± 1/16	14 mininum	14-1/2 maximum

For each Type I hanger, four spring clips shall be riveted, clinched, or welded to the frame to securely grip film with square or round (3/8-inch radius) corners, not less than 3/16 inch from the nearest film edge. The two bottom clips shall be attached to the frame. The two top clips shall be attached to springs of sufficient tension to hold the film uniformly taut without warping the film or without pulling the film out of the clips in processing. The design of the clips shall be such that trapping of solution is minimized. The overall thickness of the clip shall be no greater than 1/2 inch. All joints shall be securely riveted, spot-welded, or clinched. All welds shall be ground smooth with an iron-free grinding wheel. The size of the hanger and the name or registered trademark of the manufacturer shall be stamped on the top bar of all hangers.

Type II dental X-ray film hanger, shall be designed to securely hold 14 dental X-ray films, 7 on either side of a vertically placed supporting hanger. It shall consist of a supporting hanger, and 7 pairs of film clips. Film clip pairs shall be equally spaced on the supporting hanger with a minimum of 1-9/16 inch between horizontal centerlines. The horizontal centerline of the top pair of film clips shall be at least 3-1/4 inches below the top of the hanger and the horizontal centerline of the lowest pair of film clips shall be at least 7/8 inch from the bottom of the hanger. Unless otherwise specified, a generous tolerance of ± 1/32 inch (non-cumulative) shall apply to each dimension of the Type II hanger. The overall height of the hanger shall not exceed 14 inches; the overall width of the hanger shall not exceed 2-3/8 inches. Supporting hanger shall consist of a straight bar or strip of steel 0.031 inch thick, with commercial tolerances, curved at one end to permit suspension over the edge of a developing tank. The film clips shall be firmly and securely fastened in pairs, transversely, to the supporting hanger by welding to provide 7 clips on each side of the hanger. Each film clip shall have a maximum bite of 0.118 inch. Dental X-ray film hanger shall have three removable white matte finish tabs that will slide onto the supporting hanger. Dental X-ray film hangers shall be legibly and permanently marked with the name or registered trademark of the manufacturer.

Passivation. The hanger shall be treated to restore corrosion-resisting properties after fabrication. As an alternate to passivation, the hangers may be electropolished.

Workmanship. The hangers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). Each hanger, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6525-00-603-1760
6525-00-603-1765
6525-00-603-1775
6525-00-603-1795

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order). Purchaser should specify type of hanger and size of X-ray film to be accommodated.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

05138

Project No. 6525-1157

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.