A-A-50973
12 June 1985

COMMERCIAL ITEM DESCRIPTION

WHEEL, ABRASIVE, (Silicon Carbide, Impregnated Rubber, Course, Square Edge)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a silicon carbide abrasive wheel suitable for use in trimming dental acrylic and base plates.

Salient characteristics:

The silicon carbide abrasive wheel shall be suitable for use in trimming denture resins and base plates when mounted in a chuck on a dental grinding and polishing machine. The abrasive shall be silicon carbide consisting of 9 parts of no. 46 grit to 1 part of no. 36 grit. The bonding material shall be age-resistant synthetic rubber. The wheel shall have an open structure. The wheel shall be square edge in shape with dimensions of 3 inches ± 1/16 inch in diameter, and 3/8 inch ± 1/32 inch in thickness with a 1/4 inch ± 1/64 inch arbor hole centrally located to give proper balance. The wheel shall have a runout of less than 0.0015 inch and sides shall be flat within 0.005 inch. The wheel shall be grayish-green in color. Each wheel shall be labeled in a legible manner with manufacturer's name or trademark, diameter, thickness of wheel, and maximum operating speed (3800 surface feet per minute).

Workmanship. The wheel shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One wheel, as specified, constitutes one unit.

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one wheel. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6520-00-854-2953

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY REVIEWING ACTIVITY

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

Project No. 6520-1805

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.