A-A-50974
12 June 1985

COMMERCIAL ITEM DESCRIPTION

METAL FOIL (Tin)

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification, QQ-T-00276A (DSA-DM).

This commercial item description covers tin foil suitable for dental laboratory use in the construction of dentures.

Salient characteristics:

The tin foil shall contain not less than 99 percent tin. The tin foil shall be furnished in a roll not less than one-half pound. Dimensions shall be 6 + 1/32 inches wide with a thickness of 0.001 inch having a plus tolerance of 0.0003 inch. The spool on which the foil is wound shall be fabricated from cardboard, plastic, or other suitable material. The spool shall be of standard commercial design. As an alternate design, the spool shall have sidewalls to protect the edges of the roll of foil.

Finish. The tin foil shall have a smooth, bright polished finish.

Workmanship. The tin foil shall be free from holes, breaks, folds or wrinkles and other defects which may detract from its appearance or impair its serviceability. The foil shall be furnished clean, and free from dirt, oil, grease, wax, or other particulate matter.

Unit. Roll (RO). One roll containing one-half pound of foil, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Inspections. Inspection, as used here, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. The unit of product for inspection purposes shall be be one roll of tin foil. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-00-531-2050

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1816

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:49Z
Check the source to verify that this is the current version before use.