INCH-POUND

A-A-54987
14 October 1994

# COMMERCIAL ITEM DESCRIPTION

## TABLE, OPERATING, HOSPITAL

(Pedestal)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-T-37702, which is cancelled.

This Commercial Item Description covers a manual major surgical operating table.

Salient characteristics:

The operating table shall consist of a mobile base and pedestal, a superstructure with positioning and control mechanisms for the articulating tabletop, an articulating tabletop, and various accessories suitable for major surgical procedures.

Mobile base and pedestal. The mobile base shall be of sufficient weight and size to assure that no movement of the table shall occur in any position when supporting a 300 pound patient. The base shall be covered with corrosion-resisting steel. The base shall be equipped with a patient-grounding receptacle in accordance with National Fire Protection Association (NFPA) 99. The base shall be fitted with not less than four, 3 inch (minimum) diameter, swivel casters, and at least four floor locks which permit compensation for uneven operating room floor surfaces. Casters and floor locks shall be inboard of the periphery of the base in order to prohibit interference with personnel movement around the base. A foot pedal, located at the head end of the base, shall control the shift between the caster-supported position and the floor lock-supported positions. With the floor locks in the down position, there shall be at least a 1/2 inch clearance between the bottom of the base and the floor.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The base with the pedestal shall at all times, and in all positions, completely enclose the pedal-activated hydraulic elevating mechanism. The elevating mechanism shall be enclosed by an inverted telescoping shroud. The bottom section of this shroud shall be fitted and welded to the base cover, and ground smooth, in order to prohibit entrance of liquids. As an alternate to welding, a tightly fitted, durable gasket seal that is impervious to all liquids commonly encountered in the operating room, shall be acceptable. The pump unit shall be a completely sealed system, with the oil reservoir prefilled by the manufacturer. The foot pedal for the elevating mechanism shall be located at the head-end of the base for convenient use by the anesthetist. The unit shall be capable of being adjusted to compensate for wear. The foot pedal shall have a "dead space" between the bottom of the lifting stroke and the start of the lowering position. This shall make it necessary to apply additional pressure to the pedal, after a lifting stroke, to move the pedal to the position for lowering the tabletop. The tabletop shall raise and lower smoothly without jarring the patient. The pedestal shall prevent rotation of the tabletop with respect to the base. The base shall be so designed that it will allow the blades of a Mayo stand to fit underneath the components of the base, including the hydraulic system at its lowest position.

Superstructure. The superstructure shall be located under, and bearing-mounted to, the seat section of the tabletop. It shall enclose the operating mechanisms for the articulating tabletop sections. Gears and worms shall be machine cut from suitable materials and all load-carrying shafts shall be mounted in antifriction bearings. The superstructure shall carry, at the head end, the hand cranks for controlling the articulating tabletop sections. There shall not be more than three, or less than two, hand cranks to control the movements of the tabletop. The leftmost hand crank, when facing the head end of the superstructure, shall control only the Trendelenberg and reverse Trendelenberg positioning of the tabletop. There shall also be an indicator to show the degree of Trendelenberg, and an indicator showing degrees of lateral tilt.

The second crank shall have a function selector associated with it. The function selector shall allow the crank to control the movements of the leg section, the seat section, the back section, and flex between the back and seat sections, as desired. The kidney elevator shall be controlled by either the second crank in conjunction with the function selector, or by a third crank.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Articulating tabletop. The articulating tabletop shall consist of a leg, seat, and a back section, a kidney elevator located between the seat and back sections, and a removable headrest. The width of each component of the tabletop, excluding the headrest, shall be at least 20 inches. The length of the tabletop, with the headrest adjusted to a 1 inch clearance space between it and the back section, and excluding the footrest, shall be a minimum of 72 inches.

Leg, seat and back section. The leg, seat, and back section shall consist of a rigid, corrosion-resistant steel framework with a corrosion-resistant steel sheet cover, or cast aluminum, covered with stainless steel sheeting. The seat and back section shall be connected, and designed so that they will operate together in simultaneous motion to provide flex and reflex. Each section shall have a corrosion-resisting steel siderail running the full length of each side for attaching accessories. The head end of the leg section, and the foot end of the seat section, shall be provided with a perineal cut-out to allow use of the transurethral tray, without restricting the movement of the leg section of the tabletop. Velcro tape strips shall be applied to the longitudinal centerline of leg, seat, and back sections. The velcro strips shall be located to make with the tape strips on table top pads, and hold them securely without other types of fasteners. As an alternate to velcro strips, snaps located on the x-ray top and the table for securing the table pads to the tabletop, shall be acceptable.

Kidney elevator. The kidney elevator shall be constructed of corrosion-resisting steel, and shall adjust to 4-1/2 inches above the tabletop, as a minimum. When in the lowest position, the top of the kidney elevator shall protrude no more than 3/8 inch above the tabletop.

Headrest. The removable headrest shall be rectangular in shape, and shall be constructed of either corrosion-resisting steel, or cast aluminum. The headrest shall be at least 15-3/4 inches wide at the table end, and at least 12-1/4 inches wide at the opposite end. The headrest shall be capable of being adjusted and locked into various angles, and shall extend horizontally from the head end of the back section from 1 to 6 inches. The headrest shall raise 90 degrees above the horizontal, and lower 45 degrees below the horizontal, as a minimum.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the current version before use.

The locking mechanism for controlling the angulation of the headrest shall be of the ratchet and locking bar type. The locking surface of each ratchet tooth shall lie in a radial plane extending through the central axis of the ratchet, permitting the free end of the headrest to be raised without activating the locking bar, but requiring release of the locking bar before the free end of the headrest can be lowered. The headrest shall also be capable of being attached to the foot end of, and at a 90 degree angle to, the leg section to act as a footrest.

Accessories. Each operating table shall be provided with the following removable accessories:

* One four-section x-ray penetrable top
* Three foam rubber pads
* One anesthetist's screen
* One body restraint strap
* Two lateral body supports
* One pair of lithotomy leg holders
* One pair of padded shoulder supports
* One pair of wristlets with adjustable holders
* One intravenous arm board
* One footrest
* Support sockets as required

X-ray penetrable top. The leg, seat, back, and headrest sections of the tabletop shall each be provided with a removable x-ray penetrable top which covers the full length and width of the section. Each top shall be capable of being added to, or removed from, the table without the use of tools. Each tabletop section shall be certified and shall be plainly labeled as to being in conformance with the application requirement of the Bureau of Radiological Health. The space between the x-ray penetrable top and the tabletop surface shall be sufficient for the insertion, without binding, of a standard 14 by 17 inch cassette. The leg and back sections shall accept the cassette from either side, the head end, or the foot end. The seat and headrest sections shall accept the cassette from the foot end and head end, respectively.

A 14 by 17 inch cassette requires a minimum space measuring 18-3/16 inches long, 15-3/16 inches wide, and 3/4 inch high. When the width of the headrest is less than that of a 14 by 17 inch cassette, the brackets which attach the x-ray penetrable top to the headrest shall extend outward on the sides in such a manner as to permit the insertion of a



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

14 by 17 inch cassette from the head end. The x-ray penetrable tops, when in place, shall not impede the angulation or adjustment of the leg, seat, or back section. The headrest shall lower 40 degrees and raise 60 degrees with the x-ray top installed.

Table pads. Three foam rubber or fabric pads, minimum of 1-1/2 inches thick with electrically conductive rubber covers shall be supplied with each table. One pad shall completely cover the headrest section, one shall completely cover the leg section, and the third shall completely cover the back, kidney elevator, and seat sections. The pads shall have snaps or conductive velcro tape strips as described previously. The electrically conductive covers for the pads shall conform to the electrostatic requirements of NFPA 99.

Anesthetist's screen. The angulating anesthetist's screen shall be formed from aluminum rod and shall fit into the opening in the angulating section of the dual function sockets. The screen shall be capable of use either with or without the x-ray penetrable top and the foam rubber pads in place.

Body restraint strap. The restraint strap shall be fabricated from a machine washable, electrically conductive material conforming to the electrostatic requirements of NFPA 99. The strap shall have two quick-release type buckles fabricated of a chrome on nickel-plated metal, and shall be suitable for holding a 300 pound patient in place on the operating table throughout the range of table adjustments.

Lateral body supports. The two lateral body supports shall be formed and contoured to support a patient when the table is in the lateral position. The supports shall slide onto, and shall be capable of being secured at any point along the width of the kidney elevator. The supports shall be 3 $\pm$ 1/2 inches by 7 $\pm$ 3/4 inches and 3 $\pm$ 1/2 inches by 6 $\pm$ 1-1/4 inches. The ends of the supports shall be rounded.

Lithotomy leg holders. Each lithotomy leg holder shall have a strap on one end for supporting the patient's legs, and the other end shall fit into a siderail support socket. The leg holders shall be continuously adjustable about the vertical axis and shall also have a height adjustment. The leg holders shall not interfere with either the x-ray-penetrable top, or the foam rubber pads, and shall be supplied in pairs with sockets.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Shoulder supports. Each shoulder support shall consist of a contoured shoulder rest covered with a removable foam-rubber pad. The shoulder rest shall fit onto, and be capable of being secured at any point along the length of one leg of an L-shaped bracket. The other leg of the L-shaped bracket shall fit into the non-angulating vertical opening of the dual-function socket. The bracket shall be long enough so that the shoulder supports will not interfere with use of the x-ray penetrable top, or the foam rubber pads. Shoulder supports shall be supplied in pairs with dual function sockets.

Wristlet with adjustable holder. Each wristlet shall consist of a metal support bracket with a padded wrist restraint strap attached to one end. The other end of the support bracket shall either attach to and slide along the siderail, capable of being secured at any point along its length, or it shall fit and be secured into a separate siderail support socket. When a separate socket is needed, it shall be supplied with the wristlet. Wristlets shall be supplied in pairs and shall not interfere with either the x-ray penetrable top, or the foam rubber pads.

Intravenous arm board. The arm board shall either attach to the siderail, or fit into a siderail support socket. When placed in position, the armboard shall angulate not less than 2-1/2 degrees above the horizontal plane, and shall be adjustable, in approximately 10 or 15 degree increments throughout a minimum range of 100 degrees about the vertical axis. The armboard shall have a removable foam rubber of fabric pad with an electrically conductive rubber cover which conforms to the requirements of NFPA 99. The armboard pads shall be held in place by means of conductive velcro strips or snaps to permit easy application and removal for cleaning, and to prevent accidental removal. The arm board shall be capable of being used with the x-ray penetrable top and the foam rubber tabletop pads. When a separate siderail support socket is required for the arm board, it shall be supplied.

Footrest. The removable footrest shall be constructed of corrosion-resisting steel, and shall attach to the foot and of the leg section of the tabletop in the horizontal position and 90 degrees above the horizontal position. The footrest shall be at least 20 inches wide.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Support sockets. Each support socket shall be constructed of corrosion-resistant material, and shall slide, and be capable of being secured at any point, along the length of the siderail. The siderails shall be provided with integral, gravity operated locks, or equivalent, so that the support sockets will not slide off the ends of the siderail when the table is in steep Trendelenberg, or reverse Trendelenberg position. Each dual function socket shall have a locking angulating section with a minimum range of motion of 180 degrees, with an opening to accommodate the anesthetist's screen, and a non-angulating vertical section with an opening to accommodate a shoulder support. All sockets shall mate freely with their respective accessories.

Optional accessories. Each operating table shall be capable of accessories which must be standard with the manufacturer of the tables:

1. Knee-foot rest. The pair of cast aluminum knee-foot rests shall have locking controls to aid in the proper positioning of the patient's legs. The rests shall have straps with buckles with which to hold the patient's legs on the rests.

2. Transurethal tray. The transurethal tray shall attach to the table beneath the perineal cutouts on the leg and seat sections. The tray shall not interfere with the movement of the leg section. The tray shall contain a tissue-retaining screen, and shall have drainage facilities. The tray and the retaining screen shall be formed from corrosioning-resisting steel.

3. Bierhoff knee crutches. The pair of Bierhoff knee crutches shall have a vertical adjustment and shall have straps to hold the patient's legs in position.

4. Neurosurgery headrest-upright position. The neurosurgery upright position headrest shall be of the padded type. The headrest shall support only the chin and forehead with the patient upright in a modified chair position. The headrest shall be rigid and fully adjustable permitting flexing the head well forward for free unobstructed surgical approach to the occipital region without obstructing oral or nasal administration of anesthetics and without the chin rest contacting the chest, during respiratory movements.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

5. Neurosurgery headrest - prone position. The neurosurgery prone position headrest shall be of the padded type. The neurosurgery headrest shall be capable of supporting the patient's head in a prone position, fully adjustable for flexion, extension, and rotation of the cervical spine, with a minimum of 14 inches of horizontal adjustment from the head end of the back section of the table, with controls placed outside the sterile field and clear of the surgeon's knees.

6. Replacement neurosurgery pads. The two occipital and two frontal pads shall have all connecting hardware, be adjustable in every direction, and be suitable for use with the neurosurgery headrests.

7. Arm support. The adjustable arm support shall consist of two hollow aluminum tubes (over which a stockinette can be stretched) connected to a universal ball-joint support which attaches to a siderail support socket. The universal ball-joint shall allow movement in any plane and shall be capable of being locked into any position.

Minimum motion. The operating table shall conform to the following minimum motion requirements when it is immobilized on the floor locks and the tabletop (consisting of headrest, back, seat, kidney elevator, and leg sections, excluding the x-ray penetrable top and foam rubber pad) is level and 33 inches from the floor:

| Section | Minimum motion |
|---|---|
| Entire tabletop | * Trendelenberg: 20 degrees<br>* Reverse Trendelenberg: 20 degrees<br>* Lateral tilt to right: 10 degrees<br>* Lateral tilt to left: 10 degrees<br>* Lower to 30 inches above the (68 pedal strokes maximum) floor.<br>* Raise from 28 inches above the floor to 45 inches above the floor. A range of 27 to 43 inches shall also be acceptable. |
| Headrest | * Raise 90 degrees above horizontal<br>* Lower 45 degrees below handle<br>* Extend horizontally from back section 1 to 6 inches. |



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the currentversion before use.

| Section | Minimum motion - (continued) |
|---|---|
| Kidney elevator | * Raise a minimum of 5 inches above tabletop.<br>* Lower to within 3/8 inch above or level with or below the tabletop. |
| Back section | * Raise head-end 90 degrees above horizontal.<br>* Lower head-end 10 degrees below horizontal. |
| Seat section | * Raise perineal end not less than 15 degrees above horizontal.<br>* Lower perineal end not less than 20 degrees below horizontal. |
| Back and seat sections combined | * Flex to form an erect "V" (at the junction of the two sections) whose included angle is variable from 180 degrees to 145 degrees.<br>* Flex to form an inverted "V" (at the junction of the two sections) whose included angle is variable from 180 degrees to 155 degrees. |
| Leg section | * Lower approximately 105 degrees below horizontal. |
| Foot extension | * Apply to foot-end of and in the same plane as the leg section, and adjust at 90 degrees above horizontal leg section. |

Operation. The operating table shall operate without binding, clatter, or slipping. The flow locks, when engaged, shall completely immobilize the table and the foot pedal for the floor locks shall engage positively. When the table is at a 33 inch elevation and 20 degrees Trendelenberg, the manual controls shall not touch the floor, or obstruct the foot pedals.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the currentversion before use.

Headrest ratchet and bar strength. The headrest ratchet and bar-locking mechanism shall withstand, with no permanent deflection or distortion, a test impact of 400 inch-pounds and 440 inch-pounds.

Settling of elevating mechanism. The hydraulic pump unit shall settle no more than 1/16 inch in one hour and 1/8 inch in 24 hours.

Rigidity. With the table at its maximum height, there shall be no more than 5/64 inch rotation about the longitudinal axis of the table, 1/8 inch rotation about the vertical axis and 1-1/6 inch rotation about the transverse axis.

Electrical conductivity. The completely assembled operating table shall meet the conductivity requirements of NFPA 99.

Finish. All exposed corrosion-resisting steel parts shall be given a No. 4 commercial finish. All other steel parts shall either have a chrome on nickel-plated finish or a baked enamel finish. All other metal parts shall have a baked enamel finish. The enamel shall be applied in two coats. Each coat shall be baked on in accordance with the paint manufacturer's instructions. The color of the enamel shall be silver luster or dark gray. All finishes shall be clean, smooth and uniform and free from blistered, peeled, scratched and uncoated areas.

Service data. The contractor shall furnish, with each table, two copies each of a technical manual which contains completely instructions for installation, operation, and maintenance and lists of component parts.

Workmanship. The hospital operating table shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One operating table, together with all component parts and service data, as specified, constitutes one unit. Each unit shall be packaged in a suitable commercial container which shall provide adequate protection against corrosion, deterioration and damage during shipment and handling. Suitable noncorrosive cushioning material shall be used to protect all exterior finished surfaces of the operating table, as well as disassembled parts, accessories and components.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one hospital operating table.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one hospital operating table.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document.

6530-00-142-8795

NOTE 2: NFPA standards are available from the National Fire Protection Association, Batterymarch Park, Quincy, MA 02269.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6530-2260

Location: TBLOP13C.DOC



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER | 2 DOCUMENT DATE (YYMMDD) |
|---|---|---|
| | A-A-54987 | 941014 |

3 DOCUMENT TITLE

TABLE, OPERATING, HOSPITAL (Pedestal)

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)*

5 REASON FOR RECOMMENDATION

6 SUBMITTER

| a NAME (Last, First, Middle Initial) | b ORGANIZATION | |
|---|---|---|
| c ADDRESS (Include Zip Code) | d TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7 DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a NAME | b TELEPHONE (Include Area Code) | |
|---|---|---|
| DEFENSE PERSONNEL SUPPORT CENTER<br>ATTN: DPSC-MAP | (1) Commercial<br>(215)737-8431 | (2) AUTOVON<br>444-8431 |
| c ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756 2340 AUTOVON 289 2340 | |

DD Form 1426, OCT 89 *Previous editions are obsolete* 198-290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the current version before use.