A-A-0053134DM)
17 August 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

GLOVES, SURGEONS'
(Powder-Free, Sterile, Disposable)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a single type of powder-free surgeons' glove in various sizes made of halogenated rubber-latex.

Salient characteristics:

The powder-free surgeons' gloves shall be of the following sizes:

National Stock No.	Size
6515-01-251-3759	8-1/2
6515-01-252-2484	8
6515-01-253-4260	7-1/2
6515-01-254-4148	6
6515-01-254-7700	6-1/2
6515-01-261-1137	7

Shall be in accordance with ASTM D3577 except as follows:

4.1 Delete paragraph and substitute: "Shall be made of natural rubber latex that has been halogenated to make dusting powder unnecessary. The surface texture shall be such as to facilitate donning on the dry hand and yet provide a good grip on instruments."

5.2 Delete paragraph and substitute: "There shall be no powder or particulate matter on the inner or outer surfaces of the gloves when examined under the microscope at 900 to 1000X."

8.3 Delete paragraph and substitute:

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

"8.3 Airtight test. Gloves shall be inflated with air to a pressure of 20 mm ± 2 mm Hg while immersed in water at 70° to 75°F (21° to 24°C) to observe pinholes. The gloves shall be manipulated so that the hand of the tester does not remain on one spot of the glove and thus obscure possible leakage.

8.3.1 Alternate airtight test. Gloves shall be inflated with air to a pressure of 12 to 14 mm Hg. The glove is then sealed at the cuff. Examination shall be made after 15 minutes to determine the presence of a leak as shown by deflation."

8.4.4 Delete lines 1 and 2 and substitute "The thicknesses shall be as shown below:"

8.4.4 Add to paragraph: "The thicknesses shall be as follows:

Locaton	Thickness
Finger tip	0.13 to 0.23 mm
Palm	0.13 to 0.25 mm
Wrist	0.13 to 0.30 mm"

Add paragraph: "8.4.5 The gloves shall be beaded or banded at the cuff to prevent roll-down, or otherwise fabricated to prevent roll-down."

10.2.2 Line 1, delete "or wallets". Line 3, delete "wallet".

Workmanship. The surgeons' gloves shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 50 pairs of gloves of one size only, suitably labeled and designed for aseptic dispensing as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one pair of gloves.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number(s) are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-251-3759	GLOVES, SURGEONS', Rubber, Powder-Free, Size 8-1/2, Sterile, Disposable, 50 PR
6515-01-252-2484	GLOVES, SURGEONS', Rubber, Powder-Free, Size 8, Sterile, Disposable, 50 PR
6515-01-253-4260	GLOVES, SURGEONS', Rubber, Powder-Free, Size 7-1/2, Sterile, Disposable, 50 PR
6515-01-254-4148	GLOVES, SURGEONS', Rubber, Powder-Free, Size 6, Sterile, Disposable, 50 PR
6515-01-254-7700	GLOVES, SURGEONS', Rubber, Powder-Free, Size 6-1/2, Sterile, Disposable, 50 PR
6515-01-261-1137	GLOVES, SURGEONS', Rubber, Powder-Free, Size 7, Sterile, Disposable, 50 PR

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4168

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.