A-A-53109
31 August 1987

# COMMERCIAL ITEM DESCRIPTION

## LETTER SET, X-RAY FILM MARKING

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification GG-M-125.

This Commercial Item Description covers an x-ray film marking letter set suitable for use in identifying x-ray films.

Salient characteristics:

The x-ray film marking letter set shall be of the following types:

Set A - One complete alphabet A to Z, inclusive.

Set B - One each of letters R & L.

The letters shall be flat, gothic style, capital letters. The letters shall be 1/4 inch ±1/64 inch high. Height measurements shall be determined by measuring across the face of the letters.

Letters shall be .039 ± .010 inch thick and of a width suitably proportionate to their height. The letters shall be stamped or molded of lead. the total tin, arsenic, antimony, and copper content of the lead shall not exceed 13 percent.

Letters shall be completely and securely enclosed in formed or molded transparent plastic of the non-warping and slow burning type. The use of adhesives to bond the plastic to the letters, without a formed or molded plastic cover, is not acceptable. The plastic shall not split, crack, or separate from the letters when the letter is bent approximately 15° or dropped.

All letters shall be vertical within 1/64 inch from absolute vertical. From an established line, approximately 3/16 inch from the bottom of the plastic, all characters shall be within 1/32 inch from absolute horizontal.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The overall size of the plastic/lead letter combination shall be 0.625 inch ±0.010 inch high, 5/16 inch ±1/32 inch wide, and 0.060 inch ±0.010 inch thick. The letters shall be mounted within ±1/32 of the center line of the plastic.

Workmanship. The x-ray film marking letter set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One x-ray film marking letter set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be each like letter included in the set as specified in this description.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.

TABLE I. Sampling for examination.

| | Inspection level | AQL (defects per hundred units) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Total defects (Major and minor combined) | II | 4.0 |
| For dimensional examination | S-1 | 2.5 |

Examination. The letter set shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Classfication of Defects

| Classfication | Defect |
|---|---|
| Major | |
| 101 | Letters not flat, Gothic Style, capital letters. |
| 102 | Letters are not completely & securely enclosed in formed or molded transparent plastic. |
| 103 | Mounted letters do not slide easily into appropriate holder. |
| Minor | |
| 201 | Mounted letters not centered and vertical in plastic backing. |

Dimensional Examination. The letter set shall be examined for defects in dimensions. Any dimension not within the tolerances herein shall be classified as a defect.

Tests.

Bend Test.

Ten letters shall be selected at random for each set selected. Each end of the plastic encased letter shall be bent approximately 15°. The plastic shall not split, crack or separate from the letters and there shall be no other damage to the enclosed letters or the plastic enclosure. Any damage to the letter or plastic shall constitute a major defect.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the currenversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6525-00-604-7825 (Set A)
6525-00-604-7925 (Set B)

Note: Plastic encased lead letters shall slide easily in the holder of marker set, NSN 6525-00-608-0620.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1215

PGC 05133

Location: CID/LETTER/H13-24



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS:** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions - Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| A-A-53109 | Letter Set, X-Ray Film Marking |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS *(Street, City, State, ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* _____ |

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

*(TO DETACH THIS FORM, CUT ALONG THIS LINE.)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.