A-A-53123
17 August 1987

COMMERCIAL ITEM DESCRIPTION

CEMENT, ZINC OXIDE AND EUGENOL, DENTAL (Temporary)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a temporary, zinc oxide-eugenol type cement suitable for cementing temporary crowns, bridges and splints, and trial cementation of permanent restorations.

Salient characteristics:

The cement shall be type I, class 2 in accordance with the American Dental Association (ADA) Specification No. 30 for Dental Zinc Oxide-Eugenol Type Restorative Materials, with the following additions and exceptions to the requirements:

1. The accelerator paste shall contain carnauba wax, poly pale resin, and eugenol as the major ingredients. The base paste shall contain zinc oxide, lecithin, American White Mineral Oil - NF, colorant, and cornstarch, as major ingredients.

2. The cement shall have a setting time of 4 to 6 minutes at 75° Fahrenheit and 50 percent relative humidity.

3. The cement shall have a 24-hour compressive strength of 8 to 15 Megapascals (MPa).

4. The cement shall have 24-hour solubility and disintegration of 1.3 percent maximum by weight.

5. The cement shall be non-heat generating, radiopaque, non-irritating, and compatible with acrylic crowns. The cement shall contain no acids and shall not discolor.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

6. The cement shall be furnished as follows:

(1) 50 grams minimum of base paste in a collapsible tube.
(1) 15 grams minimum of accelerator paste in a collapsible tube.
(1) 25 sheet minimum mixing pad, 9.5X12 centimeters minimum.
(1) Instructions for use.

7. The shelf life of the cement shall be not less than 24 months from date of manufacture. The inspection test date shall be 24 months from date of manufacture. Not more than 4 months shall have elapsed from the date of manufacture of the cement to the date of delivery to the Government.

Workmanship. The cement shall be clean, uniform in quality and free from foreign material, hardening in the tube or any other defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing a minimum of 50 grams of base paste in a collapsible tube; a minimum of 15 grams of accelerator paste; a 9.5X12 centimeter 25 sheet mixing pad; and instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling and inspection procedures of ANSI/ADA Specification No. 30 shall apply for each lot of cement furnished.

Contractor certification. The contractor shall certify that the product offered is ADA certified, meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-120-2649	CEMENT, ZINC OXIDE AND EUGENOL, Dental, (Temporary).

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2010

49122

Location: NCR CID DISK 3

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53123	CEMENT, ZINC OXIDE AND EUGENOL, DENTAL (Temporary)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b ADDRESS *(Street City State ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.