INTERIM COMMERCIAL ITEM DESCRIPTION

CUFF SET, DUAL, TOURNIQUET, PNEUMATIC

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a dual cuff set for use with Tourniquet, Pneumatic, Surgical, NSN 6515-01-168-3792 to minimize extremity discomfort in patients receiving regional anesthesia during surgical procedures.

Salient characteristics:

Shall be a 24 inch (61 cm) long, dual cuff set with control valve assembly and tubing, suitable for use with the surgical, pneumatic tourniquet under NSN 6515-01-168-3792. Shall be designed to minimize extremity discomfort in patients receiving regional anesthesia during surgical procedures.

The dual cuff set shall consist of the following components:

1. An outer envelope fabricated of washable fabric. The inside surface, next to the patient, shall be of a soft stockinette for patient comfort. The outside shall include two separate lengths of non-slip hook and pile closure, and two ties for added security.

2. Two separate rubber bladders, approximately 24 inches (61 cm) long by 4 inches (10.2 cm) wide. Each bladder shall have a short length of tubing with a Luer lock type connector for secure connection to the control valve tubing.

3. Two separate plastic inserts approximately 24 inches (61 cm) long by 4 inches (10.2 cm) wide to maintain the shape of the cuff when deflated.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The control valve with tubing assembly shall consist of two one-way valves with common input, and three lengths of rubber tubing with Luer lock type connectors. Two lengths of tubing approximately 18 inches (45.7 cm) long shall connect the valves to the cuff bladder tubing. The tubing that connects the common input of the valves with the tourniquet tubing shall be a minimum of 36 inches (91.4 cm).

Workmanship. The cuff set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cuff set with control valve assembly and tubing, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one cuff set with control valve and tubing.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-253-9221

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4164

Location: Disk # 10

☆U.S. GOVERNMENT PRINTING OFFICE: 1987 - 505-033-75506

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.