A-A-0053146
3 February 1988

INTERIM COMMERCIAL ITEM DESCRIPTION

HOLDER, RADIOGRAPHIC FILM CASSETTE

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a mobile vertical type cassette holder suitable for holding standard radiographic film cassettes.

Salient characteristics:

The radiographic film cassette holder shall be a mobile, vertical type holder suitable for holding standard radiographic film cassettes up to 14 inches (35.6 cm) by 17 inches (43.2 cm).

The holder shall be a telescoping cassette carrier that can be positioned in either direction with a 180^0 tilt, have a 58 inch (147.0 cm) counter balanced vertical travel with maximum height of chin rest 71 inches (180.0 cm) above the floor.

The holder shall have a compact and easily maneuverable base plate with average dimensions of 15 inches wide (380 mm) and 23 inches long (580 mm). The base shall have four smooth-rolling, quiet ball bearing casters with a positive two-wheel locking system which makes the holder stable enough for patients to use it for support.

The holder shall be rigid, accurate, and corrosion-resisting steel construction. It shall be degreased, primed, and finished with smooth enamel in the manufacturer's standard color. Trimming shall be chrome or other corrosion-resistant material. The finish shall not fade, crack, or peel under normal conditions of use.

Welding shall be durable, sound in every detail, free of cracks, porosity, and other defects, and in accordance with the latest practices recommended by the American Welding Society.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

All holder material shall be high grade throughout and of sufficient strength to maintain operational alignment for the life of the equipment. The holder shall be free of burrs and sharp edges.

Each holder shall be furnished with one set of manufacturer's instructions for installation, operation, and maintenance.

Workmanship. The radiographic film cassette holder shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One holder, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one radiographic film cassette holder as specified in this description. Defects shall be those listed below:

Defects.

Holder not complete with all specified component parts.

Cassette carrier fails to slide smoothly and freely over the entire length of its travel.

Cassette carrier cannot be positioned in either direction with a 180° tilt.

Maximum height of chin rest less than 71 inches.

Component parts not free of burrs or sharpness at edges.

Locks are not present on at least two wheels.

Installation, operation, and maintenance instructions not furnished.

Finish not smooth, adherent, and free of chipped, cracked, blistered, or bare areas.

Holder does not roll smoothly and quietly.

Holder not clean and free of oil.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one radiographic film cassette holder.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-110-6591

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1230

PGC 40649

Location: CID/HOLDER/H13-24

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 505-033/90883

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.