A-A-53148A
13 July 1989
SUPERSEDING
A-A-0053148(DM)
13 November 1987

COMMERCIAL ITEM DESCRIPTION

SCISSORS, GENERAL SURGICAL
(Curved, 7 Inch, Metzenbaum, Tonsilectomy)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers blunt ended, curved scissors of the Metzenbaum design suitable for use in tonsilectomies.

Salient characteristics:

General. The definitions of the terms and components used in this specification are in accordance with ASTM Designation F921.

Material. The material of the scissors shall be Class 4 corrosion-resistant steel of composition 420 or similar, in accordance with ASTM Designation F899.

Hardness. The scissors shall have a finished hardness of Rockwell C50 to C55. The hardness shall be tested in accordance with ASTM Designation E18 or ASTM Designations E92 and E140.

Style, design and dimensions. The scissors shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20301.

Passivation. The complete scissors shall be passivated.

Finish. The finish of the scissors shall be mirror except for the blades which shall have a crocus finish.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Corrosion resistance. The scissors shall show no signs of corrosion resistance when tested in accordance with ASTM Designation F1089.

Performance test. Performance of the scissors shall be tested in accordance with the scissors performance test in ASTM Designation F1079. The test material shall be a medium rubber dam 0.007 to 0.009 inch thick (See Note 2). Cutting requirements apply to the entire length of blades, inclusive of the tips of the scissors.

Identification markings. The scissors shall bear the contractor's name or registered trademark and the word "stainless". Scissors of foreign manufacture shall bear the name of the country of origin.

Workmanship. The scissors shall be free from defects which detract from its appearance or impair its serviceability.

Production Quality Level.

Visual and dimensional examination. The scissors shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Material. The supplier shall be responsible for performing test assuring that materials are in accordance with requirements specified.

Unit. Each (EA). One scissors, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) inspection level of S-2. Defects shall be those listed below. The unit of product for examination purposes shall be one scissors.

Defects

- o Scissors not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
- o Finish of the scissors not as specified.
- o Style and design not as specified.
- o Scissors not free from excessive play or excessive stiffness during cutting action.
- o Identification markings not present, not complete, not correct or not legible.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one scissors. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-365-7100	Scissors, Curved, Metzenbaum, 7 Inch, Each

NOTE 2: The following National Stock Number covers the rubber dam cited in this specification:

NSN	ITEM IDENTIFICATION
6520-00-754-2769	Rubber Dam, Medium, 6 Inches Square

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4457

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/13780

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.