
FED. SUP CLASS
6640



MATERIAL: The mill shall have a stainless steel chamber, bowl and cutting blade. The cover shall be plastic with a stainless steel liner and a rubber gasket. The housing shall be cast aluminum with four rubber feet.

DESIGN: The overall dimensions of the mill shall be 102x121x203 cm high with the cover in place. The chamber dimensions, with the cover are 34 mm diameter and 34 mm high (approximately 75 ml). The mill shall have a universal motor, with sealed, self-lubricating bearing, rated at 20,000 rpm, which is designed for intermittent operation only. Motor shall be protected by a manual reset thermal overload circuit breaker. A push-button on-off switch shall be mounted on the front of the housing. Dust-tight felt gasket shall be between motor and grinding chamber for protecting the motor. The chamber cover shall be secured to housing by latching spring clips. The unit shall have a three wire cord and grounded plug for 120 volts, 60Hz, 75 watts power. The electrical wiring shall conform to all the requirements of the Underwriters Laboratories, Code Number UL73. Grinding chamber may be cooled by circulation of water through the coil around the chamber. Nozzles protruding from rear of mill take 1/4 inch I.D. tubing.

USE: The mill is used for grinding dry material down to 125 μm particle size. Performs best with brittle or easily shreddable material. Ground product is not uniform in size. Maximum charge approximately 50 grams.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

APPROVED 5 JAN 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		MILL, LABORATORY	MS36359
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1132

Source: https://assist.dla.mil -- Downloaded: 2026-10-23117:39Z
Check the source to verify that this is the current version before use.

FED IF CLASS
6640

Bath shall accommodate a graduated glass thermometer.

Water baths shall accommodate RACK, TEST TUBE, LABORATORY, MS36171-1 (20 tube type) and MS36171-2 (40 tube type) of the number specified herein.

All electrical components and insulations shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTES:

1. Voltage, cycle, and wattage shall be specified on date of invitations for bid.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 48619 and 95632.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

APPROVED REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		WATER BATH, ELECTRIC, SEROLOGICAL	MS36357
Procurement Specifications NONE	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1210

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:39Z
Check the source to verify that this is the current version before use.