A-A-0051793(DM)
28 January 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

HYPODERMIC INJECTION APPARATUS
(Jet, Automatic, Gas Pressure Power Source)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a hypodermic injection apparatus powered by carbon dioxide gas to produce a jet action for multiple inoculations in humans or animals.

Salient characteristics.

General. The hypodermic injection apparatus shall be used for rapidly injecting subcutaneously or intramuscularly metered quantities of vaccine from 0.05 to 1.0 ml. The apparatus shall have the following components:

One power unit
Two medicinal heads
Two sets of vial holders
One gas connection with gage
Four nozzle assemblies
One intra-dermal spacer
One long handle
One short handle
One carrying case

Characteristics. The power unit assembly shall be a gun-type unit with a non-slip handle grip and shall have a spring-loaded device to engage and disengage the medicinal head. The unit shall have 20 dosing blades, each representing 0.05 ml. Each blade shall be capable of being moved independently to the right for dosage selection. The power unit assembly shall have a trigger for initiating propulsion action of the unit and a safety catch. Each power unit shall be furnished with a long and a short handle. The long handle shall be designed to administer injections using 16 gram carbon dioxide cartridges. The short handle shall be designed to hold the unit when assembling the device to a carbon dioxide cylinder.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

The medicinal heads shall have a piercing needle with a needle guard, capable of penetrating the rubber cap of a vial. The heads shall be able to undergo standard steam sterilization and accept the spring-loaded nozzle assembly.

The vial holders shall be able to support multiple-dose vials containing at least 10 ml of medication and capable of being attached to the medicinal head. The holders shall maintain the vial in a secure manner.

The nozzle assembly shall have an 0.007 inch orifice and gas connection assembly that contains a standard cylinder adapter for connecting to a carbon dioxide cylinder valve outlet. The gas connection assembly shall have an integral nose with a captive, quick-disconnect connection to the apparatus proper which includes a 0-2000 psi integral pressure gage.

The apparatus shall contain an intra-dermal spacer which can be attached to the end of the injection head.

The unit shall be supplied with a maintenance kit with all the tools necessary to accomplish routine maintenance and minor repairs, instructions for operation and maintenance of the unit, and ten, 16 gram carbon dioxide cartridges (See Note 1).

All components, including the carbon dioxide cartridges, shall be enclosed in a carrying case.

The unit shall be capable of directly operating when connected to a filled carbon dioxide cylinder without needing a reducing valve or pressure regulator.

When subjected to the performance and jet nozzle test, the unit shall fire without failure or leakage around the orifice. The emitted strear shall be free from side spray.

All metal components which come in direct contact with the administered medication shall be fabricated from suitable corrosion-resistant steel, properly hardened and passivated.

Workmanship. The hypodermic injection apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One hypodermic injection apparatus, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Performance and jet nozzle test. Each unit shall be fired on a test stand for at least 50 shots. The unit being tested shall fire 25 1-cc shots and 25 1/2-cc shots of water consecutively at the rate of at least 1000 shots per hour.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Testing and examination of the finished injection apparatus shall be on a 100 percent basis. The unit of product for test purposes shall be one hypodermic injection apparatus.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers are either covered by or referenced in this document:

NSN	ITEM IDENTIFICATION
6515-01-126-4330	HYPODERMIC INJECTION APPARATUS, Jet, Automatic, Gas Pressure Power Source
4220-00-458-9240	CARTRIDGE, CARBON DIOXIDE, 16 Grams

NOTE 2: The following item is reported to comply with specification requirements:

MED-E-JET - Model 1-863 with long and short handle.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 47345

Project No. 6515-3977

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4906 WASHINGTON D.C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-005793 (DM)	HYPODERMIC INJECTION APPARATUS (Jet Automatic, Pressure Power Source)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 83 MAR — PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.