A-A-30104C
July 24, 1991
SUPERSEDING
A-A-30104B
July 17, 1987

COMMERCIAL ITEM DESCRIPTION

SPONGE, LAPAROTOMY, Sterile

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a sponge intended to be used as a packing during surgical procedures.

Salient characteristics. Sponge, Laparotomy, sterile, prewashed, disposable, 4 ply, stitched so that all plies are joined together along all four edges, and cut edges are not exposed, free from excess lint and loose threads. Each sponge shall have a 6 to 8 inches (15.24 cm to 20.32 cm) textile loop and an inert nontoxic radiopaque insert, approximately 1/2 inch by 2 to 4 inches (1.27 cm by 5.08 cm to 10.16 cm), securely stitched between layers of gauze at one corner of the sponge. If the loop is dyed, material shall be colorfast and nontoxic. The radiopaque insert shall be capable of withstanding normal steam sterilization after sponges are removed from their immediate container. The finished sponges shall show no tendency to bunch or creep/migrate.

Type I	Type III Absorbent Gauze (28-24 mesh), USP, or higher quality bleached gauze sponge to have a cross stitching (a straight line sewn connecting the mid points of each opposite side), and the immediate container shall be double wrapped.
Type II	Type III Absorbent Gauze, USP, except the gauze shall be all cotton and the immediate container shall be double wrapped.
Type III	Type III or Type IV Absorbent Gauze, USP.
Type IV	Meets requirements for Absorbent Gauze, USP, except for thread count and weight. Gauze shall be 11 to 13 by 14 to 16 threads per 2.54 cm (1 inch), fluff-dried and shall have excess crimp in the warp and fill directions. The requirements for gauze shall apply to the finished crimped gauze. As an alternate, the gauze shall meet the requirements of type III Absorbent Gauze, USP.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Type V	Meets requirements for Absorbent Gauze, USP, except for thread count and weight. Gauze shall be 11 to 13 by 14 to 16 threads per 2.54 cm (inch), fluff-dried, and shall have excess crimp in the warp and fill directions. The requirements for gauze shall apply to the finished crimped gauze. As an alternate, the gauze shall meet the requirements of Type III Absorbent Gauze, USP.
Size 1	12 inches by 12 inches (plus or minus 1-1/2 inches) (30.48 cm by 30.48 cm (plus or minus 4 cm)
Size 2	14 inches by 14 inches (plus or minus 1-1/2 inches (35.56 cm by 35.56 cm (plus or minus 4 cm)
Size 3	18 inches by 18 inches (plus or minus 2 inches) (45.72 cm by 45.72 cm (plus or minus 5.08 cm)
Size 4	8 inches (plus or minus 1 inch) by 36 inches (plus or minus 2 inches) (20.32 cm (plus or minus 2.54 cm) by 91.44 cm (plus or minus 5.08 cm)

Workmanship. The sponge shall be free from defects which detract from its appearance or impair its serviceability.

UNIT. Package (PG). Five sponges or one hundred sponges as specified, constitutes one unit.

For NSN 6510-01-067-2057 only, five sponges shall be supplied as an immediate container, capable of maintaining sterility of contents unless damaged or opened.

For NSN 6510-01-153-4689 only, five sponges shall be supplied in a sealed, peel-open, double wrapped, immediate container, capable of maintaining sterility of contents unless damaged or opened. One package containing twenty immediate containers of five sponges each (100 sponges) constitutes one unit.

For NSN 6510-01-164-2719 only, five sponges shall be supplied in a sealed, peel-open tray, capable of maintaining sterility of contents unless damaged or opened. One package containing twenty immediate containers of five sponges each (100 sponges) constitutes one unit.

For NSN 6510-01-164-2724 and 6510-01-160-3261 only, five sponges shall be supplied in a sealed, peel-open immediate container, capable of maintaining sterility of contents unless damaged or opened. One package containing twenty immediate containers of five sponges each (100 sponges) constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 20260-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Preservation, packaging, packing, labeling and marking - Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE	SIZE
	SPONGE, LAPAROTOMY		
6510-01-067-2057	45.72 cm by 45.72 cm, 5s (18 inches by 18 inches)	I	3
6510-01-153-4689	30.48 cm by 30.48 cm, 100s (12 inches by 12 inches)	II	1
6510-01-164-2724	20.32 CM by 91.44 cm, 100s (8 inches by 36 inches)	III	4
6510-01-164-2719	35.56 cm by 35.56 cm, 100s (14 inches by 14 inches)	IV	2
6510-01-160-3261	45.72 cm by 45.72, 100s (18 inches by 18 inches)	V	3

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Copies of United States Pharmacopedia (USP) regulations may be obtained from: USPC, Inc., 12601 Twinbrook Parkway, Rockville, MD 20852 USA.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type and size

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6510-0807

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.