A-A-51052A
30 May 1989
SUPERSEDING
A-A-51052
30 October, 1985

COMMERCIAL ITEM DESCRIPTION

TABLE, OVERBED

Mobile

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a mobile, adjustable overbed table suitable for use in medical facilities.

Salient characteristics:

The overbed table shall be an adjustable, single pedestal table supported by four casters. It shall be furnished in the following two types:

Type I - Dark wood grain color table with vanity and mirror.

Type II - Walnut color table.

Type I.

The table shall consist of a base, a telescoping pedestal containing an automatic lift mechanism for height adjustment, and a top. The table shall have a minimum adjustment range of 14 inches from not lower than 28 inches from floor level to tabletop, to not higher than 48 inches.

The thermosetting plastic laminate table top shall be of a dark wood grain color and have minimum dimensions of 14 inches by 30 inches. A vanity drawer with a mirror shall be attached to the underside of the tabletop.

The table base shall be an "H" or "U" shaped assembly. All open ends of tubing shall be closed with caps. The base shall be equipped with four casters. The casters shall be a ball bearing swivel caster, not less than 1-1/2 inches in diameter, or a 2 inch roller-type caster. The casters shall be capable

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

of turning and rolling freely, smoothly, and quietly. The casters shall be securely attached to the table to prevent accidental loss of casters.

With the exception of the tabletop and hardware, tables other than corrosion-resisting steel, or plated metal, shall be epoxy powder coated or painted with a suitable enamel after cleaning and priming. Any paint finish shall be free of scratches, blisters, or unpainted areas. Plated parts shall be free of scratches, pits, and peeled or unplated areas.

Type II.

The table shall consist of a base, a telescoping pedestal containing an automatic lift mechanism for height adjustment, and a top. The table shall have finger touch control or slight hand grip adjustment with a minimum height adjustment from 30 inches to 44 inches. The table shall lock into desired height.

The table shall have a solid one-piece top with dimensions of 30-32 inches long by 15 inches wide x 3/4 inch thick. The table top shall have a walnut color wood grain laminate plastic finish.

The table base shall be an "H", "K", or modified "U" type base. All open ends of tubing shall be closed with caps. The base shall be equipped with four 2-inch diameter non-marking swivel casters, one placed at each corner of the base. The casters shall be capable of turning and rolling freely, smoothly, and quietly. The casters shall be securely attached to the table to prevent accidental loss of casters.

With the exception of the tabletop and hardware, the table shall be chrome plated or have an epoxy powder finish. The epoxy powder finish shall be free of scratches, blisters, or unfinished areas. The chrome plated finish shall be free of scratches, pits, and peeled or unplated areas.

The table shall be packed one per box, fully assembled and permanently marked with name or trademark.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Type I and Type II.

The sum of the lateral deflections (side play) of the table shall not exceed one inch when a two pound lateral load is applied to the edge of the table within 3 inches of the described end. The table shall not deflect downward under load more than two inches, or take a permanent set, when a 100 pound load is evenly distributed on the area comprising the full width of the top and the 12 inches farthest from the pedestal and maintained for a period of five minutes.

Welding shall be durable, sound in every detail, and in accordance with the latest practices recommended by the American Welding Society. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement. All welds shall be continuous with no visible voids, pinholes, or cracks.

The table shall be clean and free from dirt or other contamination.

Workmanship. The overbed table shall be free from defects which detract from its appearance or impair its serviceability.

Unit. One table, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one overbed table as specified in this description.

Defects

Any component part is missing.

Any table component is fractured, split, punctured, dented, bowed, malformed, or otherwise impaired.

Any table part is loose, or not in proper alignment.

Plastic sheet is delaminated from tabletop.

Elevating mechanism binds; does not operate freely.

Table does not have a minimum of four casters.

Casters do not turn or roll freely, smoothly, and quietly.

Tabletop is not adjustable within specified range.

Exposed metal edges of table are not free of burrs and sharp edges.

Open ends of tubing are not closed.

Weld joints do not blend into the adjacent metal in gradual, smooth curves.

Welds are not continuous with no visible voids, pinholes, or cracks.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Defects - Continued.

Paint finish or epoxy powder coating is not free of scratches, blisters, or non-covered areas.

Plated parts are not free of scratches, pits, and peeled or unplated areas.

Table is not clean and free from dirt or other contamination.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one overbed table.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests.

Test for side play. The degree of side play shall be measured at the end of the tabletop opposite the pedestal when a 2 pound lateral load is applied to the edge of the table within 3 inches of the described end. Measurements shall be made at both maximum and minimum elevations of the tabletop. A sum of deflections on both sides of the rest position exceeding one inch shall constitute a defect.

Downward deflection and permanent set. The table shall be elevated to the maximum position. A 100-pound load shall be evenly distributed on the area comprising the full width of the top and the 12 inches farthest from the pedestal. The load shall be maintained for five minutes and the amount of deflection measured. The load shall be removed and the amount of set shall be determined. A downward deflection under load greater than 2 inches, or a permanent set to the table, shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-082-4902 (Type I)

6530-00-431-8303 (Type II)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2040

Location: CID/TABLE/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51052A	TABLE, OVERBED, Mobile

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*:
b ADDRESS *(Street, City, State, ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil – Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.