A-A-51757
25 November 1986

# COMMERCIAL ITEM DESCRIPTION

## CRICOTHYROTOMY KIT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cricothyrotomy kit used in administering emergency tracheotomy.

Salient characteristics:

The cricothyrotomy kit shall be a sterile, disposable kit suitable for administering an emergency tracheostomy through the cricothyroid space. At a minimum, the kit shall consist of the following components:

| Qty | Component |
|---|---|
| 1 | Needle housing unit with removable 13 gage needle |
| 1 | 4.5 ± 0.5 mm i.d. Airway with removable obturator. |
| 1 | 6.0 ± 0.6 mm i.d. Airway with removable obturator. |
| 1 | 7.2 ± 0.8 mm i.d. Airway with removable obturator. |
| 1 | Syringe, 3 ml, minimum |
| 2 ft. | suitable tie string |
| 1 | sterile kit container |

Needle housing unit. The needle housing unit shall have a cannula split along its length which automatically closes around the needle. The edges of the cannula shall be smoothly chamfered to form a tapered fit around the needle. The assembled needle and cannula shall extend 1 to 1-1/2 inches beyond the housing unit. After removal of the needle, the split cannula shall be expandable to allow passage of any specified airway. Also, the housing unit shall have 2 finger grips and have at least two slots for the purpose of tying the unit to the patient's neck.

Needle. The needle shall have a beveled point and shall be permanently attached to a female Luer lock adapter in accordance with ANSI/HIMA MD70.1. The needle point shall also have a removable, protective cap.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.

Syringe. The syringe shall be transparent or translucent and have a male Luer taper in accordance with ANSI/HIMA MD70.1.

Airways and obturators. Each airway shall extend 1/32 to 1/16 inch through the cannula of the needle housing unit. Each obturator shall be blunt tipped and extend 1/8 ± 1/32 inch through its respective airway.

The airways, scalpel blade, and cannula shall be fabricated of a suitable corrosion-resistant steel in accordance with ASTM F899. The needle shall be fabricated of a suitable Class IV corrosion-resistant steel in accordance with ASTM F899.

Workmanship. The cricothyrotomy kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cricothyrotomy kit, packaged sterile in a suitable pocket sized, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the currentversion before use.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number and the lot size.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be in accordance with MIL-STD-105, inspection level S-1 and acceptance number zero. Sampling for Sterility shall be in accordance with the USP. Unit of product for tests purposes shall be one cricothyrotomy kit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-131-3435 | CRICOTHYROTOMY KIT, Disposable, Sterile |

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 47367

Project No. 6515-3790



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.