A-A-51759
30 January 1987

COMMERCIAL ITEM DESCRIPTION

CART, MEDICINE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a medicine cart suitable for holding unit dose patient medication.

Salient characteristics:

The medicine cart shall contain medication drawers, a narcotics drawer, and miscellaneous supply drawers.

The medicine cart shall have the following approximate dimensions:

Height:	43-1/4 inches (109.85 cm)
Width:	22 inches (55.88 cm)
Depth:	28 inches (71.12 cm)

The medicine cart shall contain five open-framed, portable cassettes. Each cassette shall be primarily constructed of a 3 gauge and 7 gauge plastic coated wire frame. Each cassette shall contain 16 removable patient medication drawers. The medication drawers shall be supported by 11 gauge (plastic coated wires attached to the exterior frame.

Each of the medication drawers shall be approximately 2-1/8 inches (5.40 cm) high, 4 inches (10.16 cm) wide and 11 inches (27.94 cm) deep with a stop located on the rear of the drawer to allow for full extension of the drawer. An identification card slot, approximately 3 inches (7.62 cm) by 1-1/2 inches (3.81 cm), shall be located at the front of the drawer. An identification card shall be supplied for every medication drawer.

The medicine cart shall be supplied with patient medicine drawer dividers and a minimum of three divider slots. Each drawer shall contain three dividers to enable it to be divided into four compartments.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The medicine cart shall be supplied with a full extension narcotics drawer to be located at the bottom of the cart. The drawer shall have approximate dimensions of 5 inches (12.70 cm) high, 18 inches (45.72 cm) wide and 12 inches (30.48 cm) deep.

The medicine cart shall be supplied with a full extension miscellaneous supply drawer. The drawer shall have approximate dimensions of 5 inches (12.70 cm) high, 18 inches (45.72 cm) wide and 12 inches (30.48 cm) deep.

The entire contents of the medicine cart shall be enclosed and shall be provided with a locking mechanism for the patient medication drawers, as well as a separate lock and key for the narcotics drawer.

The medicine cart shall be constructed with a 1 inch square anodized aluminum tubular frame 0.063 inch (0.026 cm) thickness, and anodized aluminum panels, 0.065 inch (0.026 cm) thickness, or an equivalent construction.

The medicine cart shall have a mark resistant top work surface with a cart mounted goose-neck, or equivalent, battery-powered light. The cart shall have a push rail, or equivalent, and mark resistant corner bumpers.

The medicine cart shall be mounted on four 5 inch (12.70 cm) smooth rolling, quiet, heavy duty casters.

The medicine cart shall be equipped with a cup dispenser which shall accept cup diameters from 1 inch (2.54 cm) to 2-1/2 inches (6.35 cm) for medicine or water use.

The medicine cart shall be supplied with a mounted, appropriately sized, waste container.

All welds shall be complete with no breaks or cracks. Welds shall be continuous, ground smooth to blend with adjacent surfaces and clean with no pitting, spatter, or discoloration.

The cart finish shall be uniform and free of scratches, discoloration, and rough grind marks.

The cart shall be so constructed that all corners shall be rounded and edges smooth so that there will be no danger of cutting the hands of the operator by sharp corners, or tearing of clothing of an operator who may inadvertently brush against the item in passing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Workmanship. The medicine cart shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One medicine cart, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-166-9038

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1878

PGC 51354

Location: CID/CART-MED/H13-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4968 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51759	CART, MEDICINE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation.

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.