A-A-51765A
25 April 1988
SUPERSEDING
A-A-0051765(DM)
4 December 1986

COMMERCIAL ITEM DESCRIPTION

MASK, OXYGEN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a soft, adult size, multiple vent, multiple concentration range oxygen mask suitable for oxygen therapy.

Salient characteristics:

Construction. The complete unit shall essentially consist of a face mask, oxygen concentration selector, slip lock ring, humidity adapter, tubing and an elastic band.

The face mask shall be designed to fit firmly over the nose and under the chin of the adults. The edge of the mask shall be flexible to adapt to varied contours of the human face to ensure minimum leakage of oxygen. A soft metal strip or other suitable means shall be provided to adjust the mask at the nose position to prevent leakage into the corner of the eyes. The mask shall also contain two vents 13/16 ± 1/16 inch in diameter just below the nose position of the mask to release the expired carbon dioxide.

The oxygen concentration selector shall be able to attach firmly to the face mask on one end. The other end shall securely connect to the tubing. The oxygen concentration shall be selectable in the range of 24% to 50% by adjusting the dial on the oxygen selector. The dial shall be easily adjustable and shall stop firmly against the marking of the oxygen concentration selected. The oxygen concentration selection shall be marked at every 2% interval within the range of 24% to 30%. The marking interval shall be at 5% interval within the range of 30% to 50%.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The oxygen concentration selector may be in two pieces, each providing different range of oxygen concentration selector. All percentage marking shall be calibrated in liters per minute (LPM) and marked accordingly. All markings shall be permanent and clearly legible.

A locking ring or similar mechanism shall be able to lock the oxygen concentration setting and shall remain locked until manually changed.

The humidity adapter shall firmly attach over the oxygen selector. The humidity adapter shall be capable of receiving and mixing the high humidity to the oxygen flowing to the face mask.

The tubing shall have a length of 84 ± 1 inch. The inside diameter shall be 4 ± 1 mm. One end of the tubing shall adapt to the mask unit without slippage. The other end shall be suitable for all standard oxygen supply with 5 ± 1 mm to 7 ± 1 mm fittings.

The elastic band shall be 20 ± 1 inch long and 3/8 ± 1/8 inch wide suitable for holding the mask in place for patients with head sizes of varied dimensions. The band shall have reinforced end.

Material. The face mask and tubing shall be fabricated of pliable and transparent polyvinylchloride material.

The oxygen selector, locking ring and the humidity adapter shall be fabricated of suitable plastic. The locking ring and the humidity adapter shall be transparent.

All components shall be free of scratches, cracks, cuts and deformation. The tubing shall be kink resistant.

Workmanship. The mask shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty masks, each individually packaged in such a manner as to maintain cleanliness of the item until use, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for tests purposes shall be one mask. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest Revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-140-5333	MASK, OXYGEN, CLEAR PLASTIC, Multiple Concentration, Adult, Disposable, 50s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 52042

Project No. 6515-4271

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.