A-A-0051767(DM)
8 December 1986

# INTERIM COMMERCIAL ITEM DESCRIPTION

## PIN, GUIDE, CALIBRATED

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers calibrated guide pins which serve as reamer guides during hip fracture fixation.

Salient characteristics.

Characteristics. Pin shall be of one-piece construction, with a trocar point and round shank. Trocar point shall be in accordance with ASTM F366. Pin shall have a diameter of 3/32 inch (2.4 mm) and a length of 9 inches (23.0 cm).

Material. Guide pins shall be fabricated from cold drawn wire conforming to the composition specified in ASTM F138, Grade 2. Grain size of the raw material shall be five or finer.

Hardness. Guide pin shall have a Rockwell hardness of HRC 30-35.

Calibration. Starting at the point, pin shall be calibrated in 1/2 inch intervals. Calibration lines shall encircle the pin. The distance from the point shall be marked at each calibration line.

Marking and surface preparation. In addition to calibration lines and distance demarcation, manufacturers name (or trademark) and manufacturers catalog number shall be marked on the pin. Description of surface characteristics, methods of surface preparation, and methods of marking shall be in accordance with ASTM F86.

Finish. Guide pin shall have a bright, smooth, mirror finish.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Corrosion properties. Guide pin shall show no evidence of corrosion when subjected to the following test:

Boil. Guide pin shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After cleaning, the pin shall be boiled in distilled water in a glass beaker for 30 minutes, allowed to remain submerged for an additional 24 hours.

Workmanship. The guide pin shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing six pins, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The acceptable quality level (AQL) shall be 1.0 for an inspection level of S-2. The unit of product for test purposes shall be one pin.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-139-8025 | PIN, GUIDE, CALIBRATED, Trocar Point, Calibrated in Half Inches, 9 Inch (23 cm) Long, 3/32 Inch (2.4 mm) Diameter, 6s |

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC S9068

Project No. 6515-3958



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.