A-A-51770
15 December 1986

COMMERCIAL ITEM DESCRIPTION

BOTTLE, SCREW CAP, (Plastic)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-16586C.

This commercial item description covers a cylindrical, narrow mouth, plastic, screw-cap bottle with a capacity of four fluid ounces, and printed instructions regarding use of copper sulfate solution in treatment of phosphorous burns.

Salient Characteristics:

The screw-cap bottle shall be for use with stock-listed medical supply sets NSN 6545-00-924-4825 and NSN 6545-00-924-8125 as a container for copper sulfate solution to be used for treatment of phosphorous burns.

The screw-cap bottle assembly shall consist of a bottle with a screw cap closure and required printed instructions for use on the bottle and cap.

The bottle shall be of one piece construction, cylindrical in shape, with a narrow mouth and screw neck, fabricated from polyethylene material. The screw neck shall have a continuous exterior thread to accommodate the screw-cap, and a flat top surface at the opening which is level and free of ridges to allow a leak proof closure. The dimensions of the bottle shall be such that the overall height is greater than the diameter measured midway between the shoulder and the bottom of the bottle. The bottle shall be stable when placed on a flat and level surface.

The screw-cap shall be fabricated from rigid molded plastic or polypropylene material, sufficiently threaded with an interior thread for closure on the screw neck of the bottle. The inside of the cap shall be suitable to form a leakproof seal with the top surface of the bottle during closure. A liner, if supplied, shall be secured to the inside of the cap in such a manner that it does not fall out of the cap or so it is not off-center when in the closure position, and has a facing that is impervious to water and to a 1.0% concentration of copper sulfate solution.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

The overall height of the assembled bottle and screw-cap closure shall not exceed 4-9/16 inches.

The bottle shall have a 4.0 fluid ounce (120 ml) capacity. When filled to the bottom level of the screw neck with water at room temperature (25 degrees Centigrade ± 5 degrees) the fill of the bottle shall be 125 ml (+10, -5 ml).

The manufacturers normal commercial color for the bottle and screw-cap shall be acceptable, but shall be in contrasting color with the required printing.

The bottle and top of the screw-cap shall be permanently and legibly printed with the information specified below:

a. On the face of each bottle shall appear:

"BOTTLE FOR COPPER SULFATE SOLUTION, 4 OZ.
NSN 6530-00-149-0093"

b. On the back of each bottle shall appear:

"Directions for Use on Phosphorous Burns:

Prepare fresh copper sulfate solution. Use 1% concentration only. Wet burns with solution. Do not cover burns with wet pads. Lavage wound thoroughly with normal saline solution, or water if normal saline is not available. Coated phosphorous particles may then be removed."

c. On the top of the screw-cap shall appear:

"COPPER SULFATE SOLUTION BOTTLE"

The printing shall be clear, distinct and legible, of contrasting color to the finish on the bottle and screw-cap, and shall not be smeared or removed during normal handling.

Workmanship. The bottle and screw-cap, including printing, shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One bottle with screw-cap, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provision.

Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one bottle with screw-cap and printing.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-149-0093 BOTTLE, SCREW-CAP, Plastic, 4 oz.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air-Force-03

Preparing Activity:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1903

Location: CID, disk 5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51770	BOTTLE, SCREW CAP (Plastic)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.