A-A-51778A
10 April 1989
SUPERSEDING
A-A-0051778 (DM)
11 February 1987

COMMERCIAL ITEM DESCRIPTION

CATHETERIZATION SET, URETHRAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a catheterization set suitable for use during urethral catheterization procedures.

Salient characteristics:

General. The catheterization set shall be sterile, disposable and suitable for use during urethral catheterization procedures.

The set shall consist of the following items as a minimum:

One fenestrated drape of non-woven fabric with minimum dimensions 12 inches by 12 inches.

One plastic backed waterproof bed pad with minimum dimensions 12 inches by 12 inches.

Several balls of either cotton or rayon sufficient in quantity and size for prepping the patient.

One pair of disposable gloves of size 7-1/2 to 8-1/2.

One plastic specimen container with screw cap closure. The container shall have no less than 30 cc volume and shall come supplied with a self-adhesive label for listing patient's name, physician, applicable instructions and other patient information.

One plastic forceps with serrated jaws. The forceps shall be 5 inches ± 1/2 inch and with tips which meet when tested as specified herein.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

One disposable 10 ml syringe containing either sterile water or saline. The syringe shall be approximately marked "to inflate catheter only, not for IV administration".

One packet of sterile, water soluable lubricating jelly of between 2 and 7 grams.

One Iodophor prepping solution of no less than 10 cc volume. The solution shall be Povidone-Iodine Cleansing Solution, USP, at concentration of 10 percent.

The catheterization set shall be individually wrapped in a sequence lined, peel-open tray or completely sealed plastic pouch around a tray of two compartments, the larger of which shall be graduated to 1000 ml in 100 ml increments. Each set shall be wrapped to allow aseptic handling and maintain sterility of the contents until opened or damaged.

Forceps performance test. The instrument shall have no looseness or free play of either shank which would result in any misalignment of the tips in closing. Upon closing, tips which do not meet shall be rejected. Upon completion of the test, the instrument shall undergo repetition a second time with the same criteria.

Sterility and pyrogenicity. Components shall be sterile and have a minimum sterility assurance level of 1×10^{-6}.

Workmanship. The catheterization set shall be free from defects which detract from its appearance or impair its serviceability.

Expiration dating. Expiration dating period of the set shall not be less than 24 months. Not less than 5/6 of the expiration dating period shall remain at the time of delivery to the Government.

Unit. Package (PG). Each catheterization set shall be individually packaged in a peel-open container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 20 individually packaged sets, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Sets shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one set. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-334-3477	CATHETERIZATION SET, URETHRAL, w/10 ml Syringe, Sterile, Disposable, 20's

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51078

Project No. 6515-4538

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.