A-A-51780
12 January 1987

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE (Wound Closure)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile kit consisting of disposable instruments for use in soft tissue wound closure procedures.

Salient characteristics:

General. As a minimum, the kit shall consist of the following components:

Item	Quantity
Polybag, with twist tie, 7 inches wide by 10 inches deep minimum	1
Non-woven towel	1
Non-woven drape with a 9-10 square inch fenestation (centered)	1
Povidone-Iodine swab	1
Sponge, gauze, multi-layer, 3.5 x 3.5 inch minimum	4
Tissue forceps, fine tip	1
Tissue scissors	1
Needle holder, fine tip	1

Drape and towel. The drape and the towel shall each be 18 inches by 12 inches minimum. The drape and the towel shall be manufactured from multiple layers of facial tissue or a suitable non-woven absorbent material. The drape and the towel shall each be coated on one side with a water impermeable plastic film.

Swab. The swab shall be a cotton tipped wooden or rigid paper-type stick 3 to 6 inches in length. The cotton tip shall be wetted with Povidone-Iodine, 10 percent, USP. The swab shall be packaged in a foil-type container. As an alternate, 0.75 fluid ounces (min) of Povidone-Iodine may be packaged in a foil-type container and a dry swab stick or cotton ball provided in the kit separately. A minimum of 30 months before the expiration date of the Povidone-Iodine shall remain at the time of shipment to the government.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Instruments. The scissors, needle holder, and forceps shall be manufactured from carbon steel plated with nickel. The forceps tips may be cast inserts. Scissors, forceps, and needle holders shall meet all performance requirements specified under Performance Tests.

Tissue forceps. The forceps shall be tweezer style with serrated tips. The forceps shall be 5.00 ± .25 inches in length and shall have serrations for a distance of 0.250 ± 0.062 inches from the distal tip. The tips shall be separated 0.375 ± 0.062 inches in the open position. The tips may have an axial ridge and groove through the center of the serrations for alignment purposes. The tips shall be 0.062 ± 0.015 inches in width at the extreme distal tip.

Tissue Scissors. The scissors shall be 4.50 ± 0.25 inches in length and shall have 0.750 ± 0.125 inch blades. Scissor tips shall be sharp/sharp or sharp/blunt.

Needle holder. The needle holder shall be 5.00 ± 0.25 inches in length and shall be of the Webster design. The needle holder shall have a ratchet lock mechanism and shall have 0.50 ± 0.25 inch serrated or non-serrated jaws. The needle holder shall be capable of securely gripping size 5-0 (0.004 inch o.d.) sutures and 0.017 to 0.050 inch o.d. suture needles.

Instrument finish. The instruments finish shall be free from burrs, cracks, pitting, and sharp edges (non-cutting surfaces). The plating shall be uniform and free from chipping, blistering, or flaking.

Sterility. The kit and all components shall be sterile. Sterility shall be tested in accordance with the United States Pharmacopiea (USP).

Workmanship. Instruments shall be free from corrosion or foreign matter. All kit components shall be free from any defect which detract from their appearance or impair their serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing twenty individually packaged sterile kits, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Sampling for performance tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one kit. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Tissue forceps. As pressure is applied in closing the forceps in the normal manner, the extreme distal tips shall meet first. As the forceps are then gradually closed, the tips shall progressively close from the extreme distal tip back. The forceps shall be closed with no lateral pressure. Teeth shall mesh uniformly and completely. The tips shall align evenly with no overlap.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Tissue Scissors. Scissors shall make a 0.5 inch (min) cut when tested in accordance with the scissors performance test of Federal Specification GG-I-526. The test material shall be a size medium rubber dam, in accordance with Federal Specification ZZ-R-960. The cut shall be made using the distal 0.5 inch of the blade. Any snagging or binding of the test material shall constitute a test failure. After performing the scissors performance test, the blades shall show no signs of chipped or flaking plating.

Needle holder. The needle holder shall be closed to the first ratchet position, but not locked. Needle holders whose tips have not met shall be rejected. Clamp a 0.005 ± 0.001 inch o.d. cotton or silk suture (common thread may be substituted if dimensional and tensile strength requirements are met) in the jaws and subject the suture to a 0.45 pound force (minimum) for 30 seconds minimum. Any slippage of the suture in the jaws shall constitute a test failure. Clamp a 0.020 ± 0.004 inch o.d. suture needle (or, as an alternate, a steel wire of the same dimensions)in the needle holder jaws. The needle shall be clamped perpendicular to the axis of the needle holder such that a portion of the needle is exposed on both sides of the jaws. Apply a one pound force such that the force is distributed evenly on the metal protruding on either side of the jaws. Any slippage of the needle in the jaws resulting in a displacement of the needle (i.e. side to side twist is excluded) shall be a test failure.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6530-01-153-4888	SURGICAL PACK, DISPOSABLE, Wound Closure, 20s

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC-54676

Project No. 6515-3611

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.