A-A-0051781(DM)
12 Feburary 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

HOOK, TISSUE, DOUBLE PRONG (Joseph)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a double-prong Joseph tissue hook for use during surgical procedures.

Salient characteristics:

General. Shall be a sharp, Joseph, double-prong tissue hook designed for holding, elevating or exerting traction on a tissue during surgical procedures.

Material. The tissue hook shall be fabricated from Class 4 martensitic corrosion-resistant steel in accordance with ASTM F899.

Hardness. The tissue hook shall have a Rockwell hardness of C-48 to C-53 when tested for hardness in accordance with ASTM E18, or ASTM E92 and E140.

Style and design. The tissue hook shall have a sharp double-prong hook with the style and dimensions in accordance with Figure 1. The size of each hook is designated by the distance between the two points. The handle shall provide a secure grip. The handle shall be knurled or fluted.

Finish. The tissue hook shall be satin or mirror in accordance with Federal Specification GG-I-526.

Passivation. Materials used in the fabrication of the tissue hook shall be passivated.

Marking. The tissue hook shall be marked with the name or registered trademark of the manufacturer and the work "stainless". If the tissue hook is of foreign manufacture, the instrument shall bear the name of the country of manufacture.

Boil test. The tissue hook shall be tested in accordance with Federal Specification GG-I-526.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Copper sulfate test. The tissue hook shall be tested in accordance with Federal Specification GG-I-526.

Workmanship. The tissue hook shall be processed in such a manner as to be uniform and free from defects that will affect life, serviceability, or appearance.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that the design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (Ea). One hook, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tissue hook. AQL shall be 1.0 percent defective. Inspection level shall be S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-110-1373	Hook, Tissue, Double-Prong, Joseph, 5 mm
6515-01-110-1378	Hook, Tissue, Double-Prong, Joseph, 10 mm

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION
6515-01-110-1379	Hook, Tissue, Double-Prong, Joseph, 7 mm
6515-01-110-1380	Hook, Tissue, Double-Prong, Joseph, 2 mm

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3835

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.



NOTE:
FRACTIONAL TOLERANCES ± 1/32
UNLESS OTHERWISE SPECIFIED.

ALL DIMENTIONS ARE IN INCHES
UNLESS OTHERWISE SPECIFIED.

SIZE	NSN	A	B
5 MM	6515-01-110-1373	1/4	3/16
10MM	6515-01-110-1378	7/16	1/4
7 MM	6515-01-110-1379	11/32	7/32
2 MM	6515-01-110-1380	11/64	3/16

FIGURE 1 HOOK, TISSUE, DOUBLE PRONG, JOSEPH

U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-036/70813

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.