L-P-525a

FEBRUARY 10, 1958

SUPERSEDING
Int. Fed. Spec. L–P–00525 (GSA–FSS)
March 8, 1957, and
Fed. Spec. L–C–201
January 8, 1947

FEDERAL SPECIFICATION

PLASTIC STRIP, DENTAL MATRIX

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.**—This specification covers plastic strips suitable for use as matrices in cement restorations.

1.2 Classification.

1.2.1 *Sizes.*—Plastic strips shall be of the following sizes, as specified (see 6.2).

Size 1.—5 inches by ⅜ inch by 0.002 inch.

Size 2.—5 inches by ½ inch by 0.003 inch.

Size 3.—5¾ inches by ½ inch by 0.005 inch.

Size 4.—45 feet by ⅜ inch by 0.003 inch.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

GG–D–441—Dispenser; Strip, Dental.
PPP–B–566 — Boxes, Folding, Paperboard.
PPP–B–621 — Boxes, Wood, Nailed and Lock-Corner.
PPP–B–676 — Boxes, Set-Up, Paperboard.

Federal Standards:

Fed. Std. 102—Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL–L–10547 — Liners, Case, Waterproof.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129 — Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:10Z
Check the source to verify that this is the current version before use.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.—The strips shall be of clear cellulose-acetate base and sufficiently rigid to serve the intended purpose.

3.2 Dimensions. — Dimensions shall be as shown in 1.2.1.

3.2.1 *Sizes 1, 2, and 3.*—Sizes 1, 2, and 3 shall be furnished in individual strips.

3.2.2 *Size 4.*—Size 4 shall be furnished in rolls not less than 45 feet long, approximately 3/8 inch wide, and shall have a thickness of 0.003 ± 0.0005 inch. The maximum outside diameter of the roll shall be 1.56 inches and the inside diameter shall be not less than 0.50 inch in order to fit the strip dispenser in accordance with the requirements of Federal Specification GG–D–441.

3.3 Workmanship.—Workmanship shall be first class throughout. The strips shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.—Sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105. In addition, samples may be requested by the procuring activity for testing at Government laboratory.

4.2 Inspection. — Inspection may be made throughout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibility for faulty workmanship or materials which may be discovered at any time prior to acceptance. Final inspection of the finished article shall be made at point of production or at point of delivery as specified (see 6.2). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 Tests.—Shall be conducted as necessary to determine conformance with the functional requirements of this specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 *Unit of issue.*—One 45 foot roll, as specified, constitutes one unit.

5.1.2 *Packaging quantities.*—The number of units indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable for the required level of protection specified in procurement document.

TABLE I.—*Packaging quantities*

Federal Stock Number	Unit package	Intermediate package	Shipping container
6520–557–8000	1 unit	24 units	576 units

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*—Each unit shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Closure shall be adequate to prevent spillage of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:10Z
Check the source to verify that this is the current version before use.

5.1.3.2 *Intermediate package.*—Intermediate package shall be a carton of appropriate size and design, or a complete wrapping with kraft paper of suitable weight. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 **Packing.**

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be constructed in accordance with Federal Specification PPP–B–621, table III.

5.2.1.2 *Case liner.*—Each level A box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.* — Strapping, when required, shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall be constructed in accordance with Federal Specification PPP–B–621, table I.

5.2.3 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 **Marking.**

5.3.1 *Unit package.*—Each unit package shall be marked with stock number, item identification, unit, quantity, and name of contractor.

5.3.2 *Intermediate package.*—Each intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.3 *Shipping container.*—Shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 Purchasers should exercise any desired options offered herein. (See 1.2, 4.2, 5.1 and 5.2.)

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 This specification covers only the types, classes, grades, sizes, etc., of the commodity as generally purchased by the Federal Government, and is not intended to include all types, etc., which are commercially available.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:10Z
Check the source to verify that this is the current version before use.

6.5 This specification covers the following item appearing in the Armed Services Medical Stock List:

6520–557–8000 — PLASTIC STRIP, DENTAL MATRIX, Celluloid Clear, 45 feet.

6.6 Transportation description.—Transportation description applicable to this item is:

Plastic articles, not otherwise indexed by name.

Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Plastic Articles not otherwise indexed by name. (Specify weight per cu. ft.)

Motor volume minimum weights 5,600; 10,000, and 12,900 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

MILITARY INTEREST:
Army—M
Navy—Md
Air Force—AF.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:10Z
Check the source to verify that this is the current version before use.