INTERIM AMENDMENT

TO

FEDERAL SPECIFICATION

TUBE, NONMETALLIC: POLYVINYL CHLORIDE (PVC), FLEXIBLE (FOR LABORATORY AND MEDICAL USE)

This interim amendment was developed by the General Service Administration, Federal Supply Service, Washington, DC 20406, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this interim amendment as a valid exception to Federal Specification L-T-790B dated April 14, 1976.

PAGE 3

Paragraph 3.10; delete and substitute the following:

"3.10 Workmanship. Tubes shall be free from blemishes and flaws which mar their appearance or affect their serviceability."

Paragraph 4.2.1; delete in its entirety.

PAGE 4

Paragraph 4.3.1; delete and substitute the following:

"4.3.1 Visual examination. Examine tube for (colorless and clearness, dimensional, and workmanship requirements. The acceptable quality level (AQL) shall be 6.5 percent defective."

Paragraphs 4.4.4, (a), (b), and (c); delete in their entirety and substitute the following:

"4.4.4 Sterilization ability test. Sterilize samples by any acceptable method in accordance with USP. At the end of the sterilization period, examine samples for conformance to 3.6."

FSC 6720

GSA DC 77-5006

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:11Z
Check the source to verify that this is the current version before use.