
MILITARY SPECIFICATION SHEET

METERS, ELECTRICAL INDICATING, PANEL TYPE, RUGGEDIZED: VOLTMETER, DC (SQUARE FLANGE, 1-1/2 INCH), STYLE 13

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The complete requirements for procuring the meters described herein shall consist of this document and the latest issue of Specification MIL-M-10304.



INCHES	MM	INCHES	MM	INCHES	MM
.003	08	1875	4 76	.656	16.66
006	.15	.250	6.35	1.000	25 40
.0150	.38	.375	9.53	1 250	31.75
112	2 84	.500	12.70	1.312	33.32
.125	3.18	.640	16.26	1.510	38.35
				1.81	45 97

NOTES:
1. Dimensions are in inches.
2. Metric equivalents (to the nearest .01 mm) are given for general information only and are based upon 1 inch = 25.4 mm.
3. Unless otherwise specified, tolerance is ±.005 (13 mm)
4. The center line of the terminals may be located .188 (4.78 mm) above or below the center line of the meter.
5. Letters and numerals for scale numerals and captions shall be .085 (2 16 mm) max and all others shall be .050 (1.27 mm) max.

(F) denotes changes

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T13:51Z
Check the source to verify that this is the current version before use.

REQUIREMENTS:

Dimensions and configuration See figure.

Weight 4 ounces maximum.

Full-scale ranges: See table I

Scale.
Length: 1 inch minimum.
Pointer deflection: 90° nominal.

Position influence: ±3 percent change for 60° rotation from normal operating position

Accuracy: ±3 percent.

(F) Overshoot 50 percent maximum.

Response time: 2 seconds maximum.

Power consumption (loss) (at end scale deflection): 1 mA ±5 percent

High temperature cycling:
±5 percent.
±3 percent permanent change.

Temperature influence ±0 75 percent change

Exposure to extreme temperatures:
±3 percent permanent change.

Overload capacity:
Sustained overload:
±1 percent temporary zero shift.
±1 percent permanent zero shift.
±2 percent permanent change.

Dielectric withstanding voltage: 1,500 volts rms.

TABLE I Full-scale ranges

Millivolts	Volts	
Zero left	Zero left	Zero center
30	2	30-0-30
50	10	50-0-50
100	20	100-0-100
200	30	
300	50	
500	100	
	200	
	300	

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T13:51Z
Check the source to verify that this is the current version before use.

Custodians:
Army - EL
Navy - EC
Air Force - 82

Review activities:
Army - EL, AV, AT, ME, WC
Navy - EC, MC
Air Force - 11
DSA - ES

User activities:
Army - MI
Navy - CG
Air Force - None

Preparing activity:
Army - EL

Agent:
DSA - ES

(Project 6625-0574-16)

☆U.S. GOVERNMENT PRINTING OFFICE 1976-703-020/352

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T13:51Z
Check the source to verify that this is the current version before use.

FOLD

Department of the Army
Headquarters
U.S.Army Electronics Command
Fort Monmouth, New Jersey 07703

POSTAGE AND FEES PAID

OFFICIAL BUSINESS

Commanding General
U.S.Army Electronics Command
ATTN: XXXXXX DRSEL-RD-TS
Fort Monmouth, New Jersey 07703

FOLD

To detach this form, cut along this line

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T13:51Z
Check the source to verify that this is the current version before use.

To detach this fc along this line

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No 22-R255

INSTRUCTIONS. This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost Comments and the return of this form will be appreciated Fold on lines on reverse side, staple in corner, and send to preparing activity Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements

SPECIFICATION MIL-M-10304/21F METERS, ELECTRICAL INDICATING, PANEL TYPE, RUGGEDIZED: VOLTMETER, DC(SQUARE FLANGE, 1½ INCH), STYLE 13

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
A GIVE PARAGRAPH NUMBER AND WORDING.

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3 IS THE SPECIFICATION RESTRICTIVE?
☐ YES ☐ NO *(If "yes", in what way?)*

4 REMARKS *(Attach any pertinent data which may be of use in improving this specification If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64 WHICH MAY BE USED

ESC-FM 1068-68

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T13:51Z
Check the source to verify that this is the current version before use.