MIL-C-26007C(USAF)
AMENDMENT 2
22 September 1971
SUPERSEDING
AMENDMENT 1
16 May 1967

Military Specification

CHAMBER, DECOMPRESSION, ALTITUDE TRAINING, 20 MAN

This amendment forms a part of Military Specification MIL-C 26007C(USAF) dated 5 April 1967

Page 2, 2.1, under standards, add: "MS-22032-1, Recharger Assembly, Portable Oxygen", and "MS33583 Tubing End, Double Flare, Standard Dimensions for."

Page 2, 2.1, under drawings Air Force add: 7027062 Foot Rest Assy - Altitude Chamber.

Page 3, 2.2: Delete "(Copies of CGS Drawings may be obtained from CGS Scientific Corporation, Route 309, Telford, Penn)" and substitute: "(Copies of CGS Drawings may be obtained from the Procuring Activity.)"

Page 7, 3.6.1.1, line 7, after (without control panels installed): change "or" to "and". Line 8, after plumbing change "and" to "or" and after component parts, add: "of the chamber".

Page 7, 3.6.1.3, last sentence: Delete and substitute: "A rubber strike pad shall be installed on the chamber or a rubber covering provided on the cam locks to absorb the noise usually made when the cam locks are released during evacuation."

Page 8, 3.6.3, line 8: Delete "vacuum reservoir" and substitute "filter".

Page 10, 3.6.3.1: Delete the entire paragraph.

Page 10, 3.6.4, at the end of paragraph add: "Four recharger hoses shall be installed in the chamber in accordance with MS22032-1. The hoses shall be connected to the oxygen line and located on the bulkheads near the corners. The lock chamber shall be provided with one recharger hose assembly in accordance with MS22032-1."

Page 10, 3.6.4.1: Change to paragraph 3.6.4.2 and delete the last sentence. Add new paragraph 3.6.4.1 as follows: "3.6.4.1 Emergency oxygen system. A one-cylinder emergency oxygen manifold shall be installed adjacent to the lock operators panel and connected to the oxygen delivery line close to the point of chamber penetration."

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T07:45Z
Check the source to verify that this is the current version before use.

MIL-C-26007 C(USAF)
AMENDMENT 2

Page 10, 3.6.4.2, after the first sentence add: "All flared tubing shall be double flared in accordance with MS33583."

Page 11, 3.6.5, before the last sentence add: "Motors and switches used inside the chamber shall be an explosion-proof type."

Page 11, 3.6.7, after sentence 5 add: "Plastic covers used on fluorescent light fixtures shall be frosted." at the end of the paragraph add: "A minimum of six emergency light assemblies shall be installed in the main chamber. They shall be installed flush with the ceiling and equally spaced along each side of the main chamber. A minimum of two emergency lights shall be installed in the lock chamber and shall be mounted in opposite corners flush with the ceiling."

Page 14, 3.6.11, at the end of the paragraph, add: "All operators and observers positions shall have foot rests in accordance with Af Dwg 7027062. Foot switches for the intercom shall be installed in the foot rests at the positions where they are used."

Page 14, add new paragraph 3.6.12:
"3.6.12 Grab rails. Grab rails shall be installed on the outside of the chamber above all windows. The rail shall be positioned so that it will be within the reach of a person standing on the floor. The rail shall be used to assist personnel when they are mounting the observer and operators chairs."

Custodian:
Air Force - 70

Review activities:
Air Force -03,-11

Preparing activity:
Air Force -70

Project No 6930-F078

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T07:45Z
Check the source to verify that this is the current version before use.