A-A-53018
22 June 1987

COMMERCIAL ITEM DESCRIPTION

BUFFER SOLUTION, STANDARD (pH)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers color coded standard buffer pH4, 7 and 10 solutions used as calibration references for a hydrogen ion test meter.

Salient characteristics:

Shall be pH buffer solutions for use as calibration standards for a hydrogen ion test meter.

Shall be supplied as color coded solutions in suitable plastic containers with sufficient optical clarity to identify the coloring of the liquid content. Each container shall include a colored label that matches the color code of the liquid content. A dispenser type closure cap shall be supplied with each bottle.

Shall be supplied with color coding, ingredients and tolerances conforming to table I.

TABLE I.

| pH | Ingredients | Color coding | pH tolerance at 25 degrees (C) |
|---|---|---|---|
| 4.00 | Potassium Biphthalate. | Red | ± 0.02 |
| 7.00 | Potassium Phosphate Monobasic. Sodium Hydroxide. | Yellow | ± 0.02 |
| 10.00 | Potassium Carbonate. Potassium Borate. Potassium Hydroxide. | Blue | ± 0.02 |

FSC 6550

AMSC N/A

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the currentversion before use.

Shall have pH value for each solution included on a label that is affixed to the individual bottle. The indicated pH shall be traceable to the National Bureau of Standards.

Workmanship. The buffer solutions shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Bottle (BT). One bottle containing 500 ml of buffer solution, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspection and testing. Sampling for inspection and testing shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

| | |
|---|---|
| 6550-01-142-5554 | BUFFER SOLUTION, STANDARD, pH 7, 500 ml. |
| 6550-01-142-5555 | BUFFER SOLUTION, STANDARD, pH 10, 500 ml. |
| 6550-01-142-8201 | BUFFER SOLUTION, STANDARD, pH 4, 500 ml. |

Ordering data. (pH required, intermediate/exterior package quantities and labeling
and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6550-0059

Location: Disk # 3



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.