A-A-0053023(DM)
13 July 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

INCUBATOR, INFANT

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers an infant incubator for use in providing a temperature and humidity regulated environment for premature or small-for-date neonates.

Salient characteristics:

The infant incubator shall be a portable, solid state controlled unit.

The incubator shall have two modes of temperature sensing and control. When in the air or manual mode the incubator shall sense air or interior wall temperature and modulate heat input until a desired setpoint is achieved. When in the skin mode the incubator shall sense the infants temperature via a skin electrode and modulate heat input until a desired setpoint is achieved.

The minimum temperature range for the air mode shall be 28°C to 37.5°C, ± 0.4°C. The minimum temperature range for the skin mode shall be 35°C to 37.5°C, ± 0.2°C. The incubator shall have skin, air, and setpoint digital temperature readout.

The incubator shall have a water reservoir humidifier capable of raising the infant environment humidity to approximately 50 percent.

The incubator shall have a double wall hood. Fresh air supply shall be 35 L/M minimum. Fresh air shall be filtered. Air velocity over the infant mattress shall not exceed 10 CM/S.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

The incubator shall perform a temperature control system self-check upon power-up. The incubator shall alarm upon the following conditions or events:

High Temperature
Patient Temperature 1°C from skin mode setpoint
Sensor Failure
Power Failure

The incubator shall have a backup thermostat to insure high temperature alarm if the control thermostat fails.

The incubator shall operate from 115V, 60 Hz power, or as specified.

The incubator shall not exceed 143 cm in height, 70 cm in depth, and 116 cm in width. The infant bed area shall be at least 62 cm by 36 cm. The incubator shall weigh no more than 100 KG.

Incubator maximum chassis source current shall not exceed 100 microamperes rms for both ground open and ground intact conditions.

Workmanship. The infant incubator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One infant incubator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for testing and examination. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table I as specified herein. The unit of product for test purposes shall be one infant incubator.

TABLE I. Sampling for testing and examination.

	Inspection level	AQL (defects per hundred units)
Major defects	III	0.4
Minor defects	III	0.4

Tests. All tests shall be performed on a first article sample in accordance with table II.

Table II. Tests.

1. Safety
The infant incubator maximum chassis source current shall not exceed 100 microamperes rms for both ground open and ground intact conditions. Test equipment shall be a leakage current tester/safety analyzer with the capability to accurately measure chassis source current.
2. Performance.
The manufacturer shall demonstrate to the Government via mutually acceptable test methods compliance with the following specifications and performance criteria:
An air sensing mode. A skin sensing mode. An air mode temperature range accuracy of $\pm$ 0.4°C minimum. A skin mode temperature range accuracy of $\pm$ 0.2°C minimum. An ability to raise the infant environment humidity to approximately 50 percent. A fresh air supply of 35 L/M minimum. Control system shelf-check upon power-up.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the currentversion before use.

The following alarms:

High temperature
Patient temper 1°C from skin mode setpoint
Sensor failure
Power failure

Examination. each infant incubator shall be examined to determine compliance with the classification of defects in table III.

TABLE III. Classification of Defects.

Classfication	Defect
Major	
101	Control(s) non-functional.
102	Indicator non-functional.
Minor	
201	Identification marking(s) not complete, permanent, and legible.
202	Finish not free of chips, cracks, scratches, and other appearance defects.
203	Any component part not secured in place.

Examination shall not be limited to the classified possible defects listed herein.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the currentversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-299-8299

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

A-A-0053023(DM)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1967

PGC

Location: CID/INCUBATOR/H17-28

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.