A-A-53030
29 June 1987

# COMMERCIAL ITEM DESCRIPTION

## OPHTHALMOSCOPE, BINOCULAR, INDIRECT

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers a binocular indirect ophthalmoscope that is worn on the head and is used for stereoscopic and extreme fundus examinations.

Salient Characteristics.

Shall be a binocular indirect ophthalmoscope comprised of a lamp housing, binocular body, head band, transformer, three aspheric viewing lenses (14D, 20D and 28D), scleral depressor, recording charts, carrying case, teaching mirror, three replacement lamp bulbs (6.5 volts, 18 watts), two instruction manuals, and two parts manuals.

Lamp Housing - Shall be sturdy, light weight and properly ventilated for adequate heat dissipation. Shall be equipped with a heat absorbing glass filter used to absorb infrared transmissions, a condenser lens, a diffuser lens and two removable filters: one red-free filter for critical examinations of blood vessels; one ultraviolet filter used for fluorescent techniques. Shall provide for adjustment of the filament in two coordinates: Centering for optimum positioning of the filament; raised or lowered to vary the size of the spot to fill the aperture of various viewing lenses. Shall have an adjustable reflecting mirror to minimize reflections and increase effective viewing with more patients. Mirror shall be easily accessible for cleaning. Lamp socket shall be removable for easy lamp bulb replacement.

Binocular Body - Shall have an adjustable pupillary distance that can be adjusted to conform to the examiner.

Head Band - Shall be multi-adjustable with self locking, ratchet type adjusters and fitted with an aluminum frontal piece for mounting of the binocular indirect ophthalmoscope. The instrument, when mounted, shall move out of the way of the examiner without removal of the head band or disconnecting the ophthalmoscope from the head band. Shall be padded at the front, rear and top to ensure the wearers comfort.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Transformer - Shall be a reduction type for use on 110-120 volt AC, 60 cycles. Shall have at least five variable secondary voltages ranging from 2.5 to 7.5 volts and shall be a separate unit from the ophthalmoscope.

Lenses - Shall be supplied with three aspheric viewing lenses with powers of 14, 20 and 28 diopters. Shall be coated to keep reflections to a minimum and provide distortion-free viewing. Each shall be mounted in their own plastic ring used to hold the lens during examinations.

Scleral Depressor - Shall be a thimble type scleral depressor made to be worn on the finger. Shall be made of an appropriate material suitable for autoclaving.

Case - Shall be a molded carrying case used for secure storage of the indirect ophthalmoscope and accessories.

Charts - Shall be fundus paper recording charts. The unit shall be supplied with two pads, at least 50 charts per pad, one pad for the right eye and one for the left eye.

The unit shall be provided with a teaching mirror used for dual viewing for consultation or teaching and shall attach directly to the binocular body for right or left side viewing.

Workmanship. The binocular indirect ophthalmoscope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One ophthalmoscope, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the items.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| Stock Number | Item Identification |
|---|---|
| 6540-00-908-3043 | OPHTHALMOSCOPE, BINOCULAR, INDIRECT |

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6540-1179

PGC 59232

Location: DISK # 3



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

---

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

**BUSINESS REPLY MAIL**
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1 DOCUMENT NUMBER | 2 DOCUMENT TITLE |
|---|---|
| A-A-53030 | OPHTHALMOSCOPE, BINOCULAR, INDIRECT |

| 3a NAME OF SUBMITTING ORGANIZATION | 4 TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b ADDRESS *(Street City State ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* |

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

| 7a NAME OF SUBMITTER *(Last First MI)* – Optional | b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c MAILING ADDRESS *(Street City State ZIP Code)* – Optional | 8 DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.