A-A-53031
29 June 1987

COMMERCIAL ITEM DESCRIPTION

PRISM BAR, OPHTHALMIC

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers two types of Berens prism bars:

Type I - Vertical Image Displacement
Type II - Horizontal Image Displacement

The prism bars are to be used for vision screen tests to measure heterophoria or heterotropia in prism diopters and for converging and diverging exercises.

Salient Characteristics.

Material. The prism bars shall be fabricated from clear acrylic plastic such as Rohm & Hass clear plexiglas "G" or equal. Shall have a refractive index of 1.49 to 1.52, haze 1 to 3% and the transmittance of white light shall not be less than 90%. A permanent marking shall be inscribed at the bottom of each prism to identify the power.

Type I Prism Bar-Shall be the vertical image displacement type and shall consist of a bar of 14 joined horizontal base prisms of 1, 2, 3, 4, 5, 6, 8, 10, 12, 14, 16, 18, 20, and 25 diopters. A 30 diopter prism shall be supplied separately.

Type II Prism Bar-Shall be the horizontal image displacement type and shall consist of a bar of 15 joined vertical base prisms of 1, 2, 4, 6, 8, 10, 12, 14, 16, 18, 20, 25, 30, 35, and 40 diopters. A 45 diopter prism shall be supplied separately.

Case. Each prism set shall be supplied with a rigid leatherette carrying case suitable to protect them from physical damage.

Workmanship. The Prism bars shall be free from defects which detract from their appearance or impair their serviceability.

AMSC N/A FSC 6540
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One prism bar, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the items.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Stock Number	Item Identification	Type
6540-00-356-6000	PRISM BAR, Ophthalmic, Vertical Image	I
6540-00-356-6050	PRISM BAR, Ophthalmic, Horizontal Image	II

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6540-1181

Location: NCR DISK #1

PGC 05257

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.