| METRIC |

A-A-53058A
1 February 1993
SUPERSEDING
A-A-53058
13 July 1987

COMMERCIAL ITEM DESCRIPTION

PIPET, BACTERIOLOGICAL (Pasteur Type)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers disposable glass Pasteur type pipets with constrictions for cotton plugging.

Salient Characteristics:

Shall be disposable non-volumetric capillary Pasteur type pipets with constrictions that allow for cotton plugging. Shall be suitable for use in bacteriological, hematology blood bank and clinical laboratory work.

Shall be of the following ASTM designated types according to size, and of the following styles for cotton plugging (see "Ordering Data"):

Type I - 150 mm long.
Type II - 230 mm long.
Style 1 - Supplied non-plugged.
Style 2 - Supplied preplugged with cotton or synthetic fibrous material.

Shall conform to the requirements in ASTM E 732 for Type I and Type II pipets with constrictions to allow for cotton plugging.

Workmanship. The Pasteur pipets shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 250 pipets of one type and style only, as specified, constitutes one unit.

AMSC\NA FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the Pasteur pipets in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall be in accordance with commercial practice.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposal shall apply.

ASTM E 732 Disposable Pasteur-Type Pipet (Application for copies should be addressed to ASTM, 1916 Race Street, Philadelphia, PA 19103).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	ASTM Designated TYPE	Specification Designated STYLE
6640-01-359-8060	PIPET, BACTERIOLOGICAL, Pasteur Type, 9 inch (230 mm), Non-plugged, Disposable, 250s	II	1
6640-01-359-8061	PIPET, BACTERIOLOGICAL, Pasteur Type, 5-3/4 inch (150 mm), Non-plugged, Disposable, 250s	I	1
6640-01-360-0903	PIPET, BACTERIOLOGICAL Pasteur Type, 5-3/4 inch (150 mm), Preplugged, Disposable, 250s	I	2

Ordering data. Size and style pipet required, unit/intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1809

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:

1. DOCUMENT NUMBER: A-A-53058A

2 DOCUMENT DATE (YYMMDD): 1 FEBRUARY 1993

3. DOCUMENT TITLE: PIPET, BACTERIOLOGICAL (Pasteur Type)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)

b. ORGANIZATION

c ADDRESS (Include Zip Code)

d TELEPHONE (Include Area Code)
(1) Commercial
(2) AUTOVON
(If applicable)

7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME
Defense Personnel Support Center
ATTN: MSE

b TELEPHONE (Include Area Code)
(1) Commercial: (215)737-2870
(2) AUTOVON: 444-2870

c ADDRESS (Include Zip Code)
2800 S. 20th Street
Philadelphia, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT.
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198, 290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.