A-A-0053331(DM)
16 February 1988

INTERIM COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL, GAUZE, 12 PLY (DOUBLE PACK)

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This commercial item description covers 12-ply, sterile sponges, packaged two sponges per envelope.

Salient characteristics:

Shall be sterile gauze sponges suitable for the treatment of injuries, for general use in the dressing of wounds and for postoperative wound dressings. The sponges shall be sterile when tested in accordance with the USP.

The sponges shall be fabricated of bleached gauze and the gauze shall meet the requirements of type VII, Absorbent Gauze, USP, except the filling thread count may be 8 to 11 per inch. The sponges shall be folded to 12-ply and shall not have any exposed raw edges and shall be free of loose threads, ravelings and foreign matter. The finished sponges shall be of the following sizes:

SIZE	DIMENSIONS (INCHES) + 1/8 INCH	UNIT PACKAGE QUANTITY
1	4 BY 8	A (1000)
2	4 BY 4	B (1200)
1	4 BY 8	C (50)
2	4 BY 4	C (50)

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Workmanship. The sponges shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

Shall be supplied 2 sponges in a heat sealed peel open type envelope. Sterility of contents shall be guaranteed unless package is damaged or opened. Twenty-five envelopes shall be packaged in a tray.

For Size 1, unit package quantity A, one package containing 20 trays (1000 sponges), as specified, constitutes one unit.

For Size 2, unit package quantity B, one package containing 24 trays (1200 sponges), as specified, constitutes one unit.

For Sizes 1 and 2, unit package quantity C, one package containing 25 envelopes (50 sponges), as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the currentversion before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be two sponges and one envelope. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE	UNIT PACKAGE QUANTITY
6510-01-151-8145	SPONGE, SURGICAL, GAUZE, 4 BY 8 INCHES, 1000s	1	A
6510-01-153-4541	SPONGE, GAUZE, TWELVE-PLY 4 BY 4 INCHES, 1200s	2	B
6510-00-119-9314	SPONGE, SURGICAL, GAUZE, 4 BY 8 INCHES, 50s	1	C
6510-00-988-3838	SPONGE, SURGICAL, GAUZE, 4 BY 4 INCHES, 50s	2	C

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 71088

Project No. 6510-0755

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the current version before use.