A-A-53331A
13 March 1989
SUPERSEDING
A-A-53331(DM)
16 December 1987

COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL, GAUZE, (12 PLY, DOUBLE PACK)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers 12-ply, sterile sponges, packaged two sponges per envelope.

Salient characteristics.

Shall be sterile gauze sponges suitable for the treatment of injuries, for general use in the dressing of wounds and for postoperative wound dressings. The sponges shall be sterile when tested in accordance with the USP.

The sponges shall be fabricated of bleached gauze and the gauze shall meet the requirements of Absorbent Gauze, USP, Type VII except the filling thread count may be 8 to 11 per inch. The sponges shall be folded to 12-ply and shall not have any exposed raw edges and shall be free of loose threads, ravelings and foreign matter. The finished sponges shall be of the following sizes:

SIZE	DIMENSIONS (INCHES) +/- 1/4 INCH	UNIT PACKAGE QUANTITY
1	4 by 8	A (1000)
2	4 by 4	B (1200)
1	4 by 8	C (50)
2	4 by 4	C (50)

Workmanship. The sponges shall be free from defects which detract from their appearance or impair their serviceability.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Unit. Package (PG).

Shall be supplied 2 sponges in a heat sealed peel open type envelope. Sterility of contents shall be guaranteed unless package is damaged or opened. Twenty-five envelopes shall be packaged in a tray.

For Size 1, Unit Package Quantity A, one package containing 20 trays (1000 sponges), as specified, constitutes one unit.

For Size 2, Unit Package Quantity B, one package containing 24 trays (1200 sponges), as specified, constitutes one unit.

For Sizes 1 and 2, Unit Package Quantity C, one package containing 25 envelopes (50 sponges), as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one sponge and one envelope. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-4, AQL 6.5 maximum percent defective except that for major raw edges (over 1/4 inch long) the AQL shall be 2.5 and for major contamination (anything larger than minute specs and obviously visible) the AQL shall be 1.0 maximum percent defective. For minor contamination (minute specs) and minor raw edges (between 1/8 and 1/4 inch long) the AQL shall be 6.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE	UNIT PACKAGE QUANTITY
6510-01-151-8145	SPONGE, SURGICAL, GAUZE, 4 by 8 inches, Two Sponges in Peel Open Type Envelope, 1000s	1	A
6510-01-153-4541	SPONGE, GAUZE, TWELVE-PLY, 4 by 4 inches, 1200s	2	B
6510-00-119-9314	SPONGE, SURGICAL, GAUZE, 4 by 8 inches, 50s	1	C
6510-00-988-3838	SPONGE, SURGICAL, GAUZE, 4 by 4 inches, 50s	2	C

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 05051

Project No. 6510-0756

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITIONS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTA
NECESSA
IF MAILE
IN THE
UNITED STA

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.