A-A-53336
1 February 1988

COMMERCIAL ITEM DESCRIPTION

KNIFE, PERIODONTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers the following double-ended knives suitable for use in periodontal surgery.

Orban No. 1 and 2

Kirkland No. 15K and 16K

Salient characteristics:

General. The knives shall be in accordance with the applicable requirements of Federal Specification GG-I-526, with the options and additional requirements specified herein.

Bid sample. Twelve bid samples, which are representative of the knives to be supplied, shall be furnished complete with protocols and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center for evaluation by the Military Medical Services. When tested by periodontists under conditions of normal usage, the knives shall be suitable for use in periodontal surgical procedures. The knives shall be well balanced, with the blade finished and contoured to produce a cutting edge around the blade which will make a sharp, accurate incision when used in dissecting and removing gingival tissue from the anterior or posterior (interproximal and lingual surfaces) regions of the mouth for Orban knives, and when used in the resection of the gingiva in all areas of the mouth for Kirkland knives.

Material.

Operative ends. The operative ends shall be class 4, type 440A, 440B or 440C stainless steel.

Handles. The handles shall be fabricated from chromium or nickel plated brass, or martensitic or ferritic stainless steel (class 4, ASTM F899) for the applicable handle design as specified on DPSC drawing 24419 and 24401 for the applicable knife.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A, C54 to C58 for type 440B, and C56 to C60 for type 440C.

Design and dimensions.

Orban, Number 1 and 2. The knife shall be a double-ended knife of Orban design, Number 1 and 2, which has thin, double-edged, spear shaped cutting blades. The design, and dimensions of the knife shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24419. Handles furnished shall conform to handle design specified on the drawing (including respective shank design). The knife shall be suitable for use in making initial incisions in dissecting gingival tissue from anterior or posterior regions of the mouth during periodontal treatment.

Kirkland, Number 15K and 16K. The knife shall be a double-ended knife of Kirkland design, Number 15K and 16K, with thin scalene-shaped broad-blades, which shall have a cutting edge around the entire periphery of the blade. The design, and dimensions of the knife shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24401. Handles furnished shall conform to handle design specified on the drawing (including respective shank design). The knife shall be suitable for use in making initial incisions in dissection of gingival tissue during periodontal treatment.

Construction. The knives shall be of three-piece construction. Handles shall be of solid or hollow construction as specified on the drawing for the respective handle. Knurled or serrated portions of handles shall be as specified on the drawing for the respective handle. The operative ends shall be securely attached to the handles. There shall be no movement, loosening, or separation between the operative ends and handles when the operative ends are subjected to an 85 pound (minimum) pull out force when tested as specified herein.

Finish.

Operative ends. The operative ends shall have a mirror or highly polished finish blended to form a sharp cutting edge. The cutting edge of the blade of the operative end shall be formed by the junction of the smoothly contoured inner and outer surfaces of the blade. The operative end (blade) shall be free of defects, such as grind marks, when viewed as specified herein. The cutting edge shall be central and uniform. The cutting edge bevel shall be smooth and continuous throughout the entire peripheral length of the blade.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

The cutting edge shall be free from micro-cracks and shall not contain nicks, feathers, burrs, or jags as specified herein, when viewed microscopically as specified herein. The cutting edge shall not contain any nick 0.004 inch or greater nor no more than two nicks 0.002 inch or greater, when measured in any direction. (A nick shall be defined as a chipped-out, broken-out, indented, or bent-out piece of metal or similar gap, indentations or projection). The cutting edge shall not contain any feathers 0.005 inch or greater, nor more than two feathers smaller than 0.005 inch in length. (A feather shall be defined as a thin, turned or curled edge not removed by the finishing process). No two nicks or feathers of the dimensions specified above shall be closer than 1/8 inch apart. Measurement of distance between defects shall be made along a straight line connecting the two defects. There shall be no burrs, jags, or defects other than those as defined herein on the cutting edge. (A burr shall be defined as a piece of metal projection not normally inherent to a smooth uniform surface). (A jag shall be defined as several small tooth-like projections, or indentations, individually smaller than a nick, 0.002 inch, and having a total continuous length of 0.015 inch or greater).

Handles and shanks. Finger grip portions of handles shall be serrated or knurled. The handles and shanks shall have a bright finish, except for the serrated or knurled portion of the handle. The serrated or knurled portion of the handles shall be free of barbs and shall provide adequate finger grip during use. The interface of the shanks and handles shall be smoothly blended with no crevices evident.

Corrosion resistance. The copper sulfate test requirements of GG-I-526 shall apply to the operative ends and martensitic and ferritic stainless steel handles. The boil test requirements of GG-I-526 shall apply to the operative ends and all handles. The salt spray test specified for brass composition in GG-I-526 shall not apply.

Identification marking. Each instrument shall be permanently and legibly marked with the instrument identification which shall appropriately identify the instrument as an Orban 1 and 2 or Kirkland 15K and 16K. All markings shall be on the smooth portion of the handle.

Workmanship. In addition to the requirements of GG-I-526, the operative ends shall be free of discoloration, pits, and rounded edges.

Unit. Each (EA). One knife, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one knife. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests, acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall include but not limited to the following:

Design of instrument not as specified.
Instrument not constructed as specified.
Operative end finish not as specified.
Operative end not free from cracks, grind marks, or burrs.
Cutting edges not sharp.

Handle finish not as specified.
Instrument not permanently and legibly marked with name or trademark of the contractor.
Instrument identification not present or properly located.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Tests. Test shall be conducted to determine compliance with specification requirements.

Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double-ended instrument.

Microscopic inspection test. The operative end (total surface of knife blade and cutting edge) shall be inspected under ten power magnification (10X) using a binocular microscope and direct (onto the top of the stage) illumination by two 15 watt fluorescent lamps. The lamps shall be positioned to provide the optimum illumination necessary for examination of the blade. The examining background shall be dark and non-reflective. The cutting edge of the blade shall be inspected on all surfaces (blade face flat and with cutting edges up).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of thos document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-01-046-7330	KNIFE, PERIODONTAL, Double-Ended, Orban, No. 1 and 2.
6520-01-046-8420	KNIFE, PERIODONTAL, Double-Ended, Kirkland, No. 15K and 16K.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2042

Location: NCR Disk 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, restrictive, loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN. DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53336	KNIFE, PERIODONTAL

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.