A-A-53348
15 February 1988

COMMERCIAL ITEM DESCRIPTION

PLASTER, DENTAL, (Modeling)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification SS-P-446D.

This commercial item description covers plaster of paris suitable for making dental models. The plaster shall be furnished in the following nominal sizes, as specified:

Size 1 - 35 pounds
Size 2 - 4 pounds

Salient Characteristics:

The plaster shall be suitable for accurately reproducing the recorded dental details of a negative impression.

Composition and physical properties. The plaster shall comply with the requirements of American Dental Association (ADA) Specification No. 25 (Dental Gypsum Products) for composition and physical properties of Type II, model plaster. Test procedures for setting time, fineness, setting expansion, compressive strength, testing consistency, and reproduction of detail shall be as specified in ADA Specification No. 25.

Net weight. The net weight of the plaster shall be a minimum of 35 pounds for size 1 and 3 1/2 pounds for size 2, when tested using a suitable balance or scale.

Delivery schedule. Not more than 3 months shall have elapsed from the date of manufacture of the plaster to the date of delivery to the Government.

Instructions. Information shall be as required by ADA Specification No. 25 and include a statement that heavy ingredients settle to the bottom, and that for consistently accurate casts, plaster should be mixed thoroughly before dispensing. In addition, two copies of the instructions shall be forwarded to the contracting officer for information purposes.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Certification. Suppliers shall furnish, at the time of award of the contract, a dated certificate and data showing that the dental plaster they propose to furnish is the same in quality and brand as that appearing on the list of certified materials for American Dental Association Specification No. 25 for Dental Gypsum Products. The plaster shall be certified for Type II of ADA Specification No. 25. Compliance with this certification requirement does not absolve the contractor from complete compliance with the requirements of this specification.

Workmanship. Workmanship shall be first class throughout. The plaster shall be free from defects which detract from its appearance or may impair its serviceability.

Unit. Can (CN). One moisture-proof recloseable container containing 35 pounds of plaster (Size 1) or 3 1/2 to 4 pounds of plaster (Size 2), together with manufacturer's recommended instructions for use, as specified, constitutes one unit.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in Tables I, II and III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table I. Unit of product for examination purposes shall be the unit of issue.

TABLE I. Sampling for Examination

Type Defect	Inspection Level	AQL (Percent Defective)
Major	II	1.0
Minor	II	2.5

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Sampling for tests. Sampling for tests shall be conducted on each lot or control number in accordance with Table II. If any test determination for a lot does not comply with the requirements specified, the lot shall be rejected.

TABLE II. Tests

Characteristic	Number of determinations per lot or batch number
Consistency	3 1/
Setting time	5 1/
Setting expansion	2 1/
Compressive strength	5 2/
Fineness	2 1/
Net weight	3 3/
Reproduction of detail	2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

1/ Each determination shall be the average of two specimens.
2/ Each determination shall be the average of five specimens.
3/ In process testing may be used for this characteristic.

Sample unit usage. Where feasible, the same unit of product shall be used for the determination of two or more test characteristics.

Examination. The plaster shall be examined to determine compliance with the requirements of this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table III.

TABLE III. Classification of defects

Categories	Defects
Major:	
101 1/	Plaster not free from foreign matter, dirt, or impurities.
102 1/	Plaster not free from caked or set lumps.
103 1/	Instructions for use inadequate, incomplete, inaccurate, or missing altogether.
Minor:	
201 1/	Plaster not uniformly white in color throughout.
202	Packaging not as specified.
203	Packing not as specified.
204	Marking incorrect, incomplete, illegible, or omitted.

1/ In process examination may be used in examination for these defects.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing, shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity. and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification	Size
6520-00-203-7235	PLASTER, DENTAL, Modeling, 35lb (15.75kg)	1
6520-00-687-7892	PLASTER, DENTAL, Modeling, 4lb (1.81kg)	2

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

A-A-53348

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2054

05393

Location: NCR Disk #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53348	PLASTER, DENTAL, (Modeling)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.