A-A-53366
15 February 1988

COMMERCIAL ITEM DESCRIPTION

CURETTES, PERIODONTAL, (Gracey)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers double-ended periodontal curettes of Gracey design.

Gracey No. 1 and 2
Gracey No. 3 and 4
Gracey No. 5 and 6
Gracey No. 7 and 8
Gracey No. 9 and 10
Gracey No. 11 and 12
Gracey No. 13 and 14

Salient characteristics:

General. The curettes shall be in accordance with applicable requirements of Federal Specification GG-I-526 with the exceptions and with the additional requirements specified herein.

Bid sample. Twelve bid samples which are respresentative of the curettes to be supplied shall be furnished complete with protocols and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center, for evaluation by the Military Medical Services. When tested by periodontists under conditions of normal usage, the curettes shall be suitable for use in periodontal surgical procedures. The curettes shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade which shall effectively accomplish subgingival curettage, remove serumal calculus from root surfaces and thoroughly polish the cementum to assist in physiological tissue reattachment. The finish of the finger grip portion of the handle shall be professionally acceptable in providing proper control and comfort during operative use.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Material.

Operative ends. The operative ends shall be Class 4, Type 440A, 440B, or 440C stainless steel.

Handles. Handle No. 1 and 3 shall be chromium or nickel plated brass. Brass shall conform to copper alloy No. 260 or No. 330, as specified in ASTM B135 for chemical composition. Handle No. 2 shall be fabricated of Class 3, Type 303, austenitic stainless steel. Handle No. 4 shall be fabricated of Class 4, Type 410 or 416, martensitic stainless steel. The corrosion tests, as specified in Federal Specification GG-I-526 shall apply as applicable.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for Type 440A, C54 to C58 for Type 440B, and C56 to C60 for Type 440C. Handle and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended.

Design and dimensions. The curettes shall be double-ended Gracey curettes. Curettes shall be furnished in the applicable number specified. The curettes shall be fine bladed and suitable for use in subgingival curettage, removal of serumal calculus from root surfaces, and for thoroughly polishing the cementum to assist in physiological tissue reattachment during periodontal treatment. The curettes shall be designed to provide push and pull action during curettage with minimum damage to tissues. The curettes shall be suitable for the purpose specified in table I. The design and dimensions of the curettes shall conform to Defense Personnel Support Center (DPSC) Drawing 24370 for the applicable number. Either handle 1,2,3, or 4 (with respective shank designs) shall be furnished.

TABLE I. Gracey numbers and purpose

Gracey Number	Purpose
1 and 2	anterior teeth
3 and 4	anterior teeth
5 and 6	bicuspids and molars
7 and 8	bicuspids and molars

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

TABLE I. Gracey numbers and purpose (cont.)

Gracey Number	Purpose
9 and 10	less accessible root surfaces of molars
11 and 12	mesio-lingual root surfaces of bicuspids and molars
13 and 14	distal root surfaces of bicuspids and molars

Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, function, fit, assembly and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

Construction. The curettes shall be of double-ended construction consisting of a handle and an operative end at each end. The operative ends shall be permanently secured to the handle. The blades of the operative ends of each curette shall face in opposite directions. Handle No. 1 and 3 shall be of hollow construction. Handle No. 2 and 4 shall be of solid construction. The interface of the shank and handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Finish. The operative ends including shank shall have a smooth mirror or smooth bright finish, except for the surfaces that form the cutting edges which shall have a finely honed finish. The operative ends (entire blade) shall have a sharp cutting edge and shall be free of defects such as rough grind marks, cracks, gouges, burrs, feathers, or other imperfections when viewed as specified herein. Honed portion of operative end shall be blended at junction of shank to form a smooth contour. Handle shall have a bright finish, except for the serrated or knurled portions. Finger grip or serrated portions of handle shall be free from barbs and shall provide adequate finger grip during use.

Corrosion resistance. Corrosion resistance tests in accordance with Federal Specification GG-I-526 shall be applicable. The copper sulfate test shall apply to operative ends and martensitic stainless steel handles. The boil test shall apply to entire curette, (operative ends and all handles). The salt spray test specified for brass composition in Federal Specification GG-I-526 shall not apply.

Identification marking. In addition to the requirements of Federal Specification GG-I-526, each curette shall be permanently and legibly marked with the word "Gracey" and the respective Gracey instrument number, as applicable. Each instrument shall be marked on the front face. All marking shall be on the smooth portion of the handle and shall be permanent and legible. Each curette shall be marked with the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless," when applicable, shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless," the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

Workmanship. In addition to the requirements of Federal Specification GG-I-526, the operative ends shall be free of pits, discoloration, and rounded edges.

Unit. Each (EA). One curette, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one curette. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests, acceptance number shall be zero (0) for all sample sizes.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Examination. Examination shall include, but not be limited to, the following defects:

Design not as specified for the applicable number curette.
Instrument not constructed as specified.
Operative end finish not as specified.
Operative end not free from rough grind marks, cracks, burrs, gouges, feathers, other imperfections that would render the curette unsuitable for use when examined under four power (4X) magnification.
Cutting edges rounded or not sharp, when examined under four power (4X) magnification.
Honed portion of operative ends not blended at junction of shank to form a smooth contour.
Operative end dimensions not as specified.
Handle finish not as specified.
Instrument not permanently and legibly marked with name or trademark of the contractor.
Instrument identification not present, properly located, or not complete.
Handle dimensions not as specified.
Curette not free from dirt, oil, compound, or foreign matter.
Handle serrations missing, not uniform or not complete.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Hardness. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double-ended instrument.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-01-005-6887/6888	CURETTES, PERIODONTAL (Gracey) No. 11,12,13,14
6520-01-006-3256/57/58/59	CURETTES, PERIODONTAL (Gracey) No. 1,2,3,4,5,6,7,8
6520-01-007-6118	CURETTES, PERIODONTAL (Gracey) No. 9 and 10

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
FDA-MPQAS
PHS

PROJECT NO. 6520-2053

Location: NCR DISK 2

05778

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53366	CURETTES, PERIODONTAL, (Gracey)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.