A A-53369
12 February 1988

COMMERCIAL ITEM DESCRIPTION

CONDUCTOR, BONE CUTTING WIRE SAW (Bailey)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a Bailey bone cutting wire saw conductor for use in introducing a Gigli bone wire saw.

Salient characteristics.

General. Shall be a Bailey conductor for use with a Gigli bone wire saw (See note 2).

Material The conductor shall be fabricated from nickel-plated carbon steel in accordance with Federal Specification GG-I-526. As alternates, the guide tip may be fabricated from Class 3 austenitic stainless steel in accordance with ASTM F899 and the conductor body may be fabricated from Class 4 martensitic stainless steel in accordance with ASTM F899

Hardness. Conductors fabricated from carbon shall have a Rockwell superfacial hardness, 45-N, 54 to 58. Conductor guide tips fabricated from Class 3 austenitic stainless steel shall have a Rockwell hardness of B75 to B90 The conductor body shall have a Rockwell hardness of C40 to C47.

Style and design The style, design and dimensions of the conductor shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20446.

Passivation Materials used in the fabrication of the conductor shall be passivated if required.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Finish. The finish of the conductor shall be mirror, satin or matte. The matte finish shall exhibit a uniform, nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Boil test. The conductor shall meet the boil test requirements as specified in Federal Specification GG-I-526.

Copper sulfate test. When the conductor is fabricated of nickel-plated carbon steel, it shall meet the copper sulfate test requirements as specified in GG-I-526.

Identification markings. The conductor shall bear the manufacturer's name or registered trademark and the word "stainless". Conductors of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The conductor shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests assuring that materials are in accordance with requirements specified.

Unit. Each (EA). One conductor, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for visual examination shall be in accordance with MIL-STD-105. The unit of product for examination shall be one conductor. Sampling shall be Level II with an AQL of 6.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for tests purposes shall be one conductor. AQL shall be 1.0 (percent defective). Inspection level shall be S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to thee producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subscontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1. The following National Stock Number is covered by this document.

NSN	ITEM IDENTIFICATION
6515-00-363-2300	CONDUCTOR, BONE CUTTING WIRE SAW, Bailey

NOTE 2. The Bailey bone cutting wire saw conductor shall be used with a Gigli bone cutting wire saw handle, NSN 6515-00-363-2400 and a 20 Inch bone cutting wire saw, NSN 6515-00-363-2700, to form a Gigli bone saw.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PBS
FDA-MPQAS

PGC 45579

Project No. 6515-4167

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53369	CONDUCTOR, BONE CUTTING WIRE SAW (Bailey)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b ADDRESS *(Street City State ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Word ing

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 62 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.