A-A-53371A
19 June 1989
SUPERSEDING
A-A-0053371 (DM)
15 February 1989

COMMERCIAL ITEM DESCRIPTION

STOP, ENDODONTIC, SILICONE

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers endodontic stops for use with endodontic files and reamers.

Salient characteristics:

The endodontic stops and their dispenser, shall be used for marking, gauging and stopping the working length of endodontic files and reamers. The stops shall be 4.0 ± 0.1 millimeters in diameter, 1.0 ± 0.1 millimeter thick, circular discs made of silicone rubber. The stops shall have a reference point which shall be a circular cutout on the diameter of the disc no greater than 0.5 millimeter in depth. The stops shall be red or of another color which sharply contrasts the teeth. A minimum of 200 stops shall be contained in a rectangular dispenser which shall have a well for removal of the stops. The dispenser shall have a spring-action mechanism for opening and closing the well. The mechanism shall release a single stop so that an endodontic instrument may pierce the stop and remove it from the well. The dispenser shall have a built-in length measurement device located on a side adjacent to the well opening. The built-in length measurement device shall consist of a metric scale and a groove where the instrument with the stop can be placed and set at the desired length. The scale shall begin at the distal end of the groove for the stop. The scale shall have a range of 0-35 mm permanently numbered every 10 millimeters with 1 millimeter line markings between the numbers. Each stop shall have a tiny puncture through its center to serve as a guide for insertion of an endodontic instrument.

Autoclavability. The stops shall show no evidence of deterioration when autoclaved as specified herein.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Workmanship. The dispenser and stops shall be free from defects which either detract from their appearance or may impair their serviceability, proper functioning, or intended use.

Unit. Package (PG). One package containing two hundred stops in dispenser, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one dispenser of 200 stops, except for dimensions 5 stops from each dispenser sampled shall be examined, and for count, no more than two units of product shall be examined.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Defects

Opening/closing mechanism on dispenser not functioning.
Dimensions not as specified.
Cutout on stop missing.
Puncture at center of stop missing.
Grooves missing or improperly placed.
Scale markings incomplete, missing or incorrect.
Dispensing mechanism does not release one stop at a time.
Number of stops in dispenser not as specified.

Sampling for tests. Sampling for tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one dispenser of 200 stops, except only 5 stops from each dispenser sampled shall be tested.

Tests. Tests shall be conducted to determine compliance with the specification requirements.

Autoclave test. Five stops from each dispenser sampled shall be autoclaved at $270^{0}F \pm 5^{0}F$ and 27 ± 1 psi pressure for a minimum of 20 minutes. After the stop is removed from the autoclave, the surface area shall be examined for any evidence of deterioration.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-01-234-6841	STOP, ENDODONTIC SILICONE, 200 stops in dispenser

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2265

LOCATION: NCR DISK # 5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53371A	STOP, ENDODONTIC, SILICONE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
3b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.