A-A-54211
13 July 1990

COMMERCIAL ITEM DESCRIPTION

CARVER, DENTAL, (Binangle, Cleoid-Discoid, No. UWD-5)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

This commercial item description covers a double-ended dental carver suitable for use in carving amalgam restorations in the plastic state, following their condensation in cavities prepared within the teeth.

Carvers shall be of either one-piece or three-piece construction.

Material.

One-piece. One-piece instruments shall be fabricated entirely of Class 4, type 440A martensitic stainless steel as specified in ASTM Standard F899, for chemical composition only.

Three-piece

Handle. The handles of the three-piece instruments shall be brass or stainless steel. Brass shall conform to copper alloy number C26000 of ASTM B135, for chemical composition only. Brass handles shall be chromium-over-nickel plated.

Operative end. The operative ends of the three-piece instruments shall be Class 4, type 440A martensitic stainless steel as specified in ASTM F899, for chemical composition only.

Hardness. After proper stress relieving, the operative ends shall have a Rockwell hardness of C52 to C56 when tested as specified in ASTM E18 or E92. Handles and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. Carvers shall be double-ended. The design and dimensions of the carvers shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24602. Carvers shall be suitable for carving amalgam restorations and wax patterns.

Construction. One piece instruments shall have an octagonal handle. The handle shall be serrated on seven faces and smooth on one face. The instrument identification shall be permanently and legibly marked on the smooth face.

Three-piece instruments shall have a hollow handle, round in cross section. The handle shall have a smooth portion and a finely serrated or diamond knurled portion. The instrument identification shall be permanently and legibly marked on the smooth portion. The shanks of the operative ends shall be permanently secured in the handle by brazing. The interface of the shank and handle shall be smoothly blended, with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force.

Each carver shall be wrapped in neutral tissue or other suitable anti-corrosive material and shall be placed in a close fitting container to prevent damage to the operative ends.

Finish. The operative ends of the carvers shall be finished to provide a sharp edge suitable for carving amalgam and wax patterns. The outer face of the cleoid and discoid ends shall have a bright finish, and the inner face shall have a fine honed finish. The shanks shall be smooth and shall have a bright polished finish. The handles shall have a bright polished finish.

Corrosion-resistance. Corrosion resistance tests in accordance with ASTM F1089 shall be applicable. The entire instrument shall meet the boil test requirement. Stainless steel parts shall meet the copper sulfate test requirement.

Identification marking. Carvers shall be permanently and legibly marked on the handle of the instrument. Marking shall include the name or registered trademark of the manufacturer or contractor, the instrument number, and the word "stainless". Instruments of foreign manufacture shall bear the name of the country of origin.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Workmanship. The carvers shall be free from coarse grind marks, cracks, burrs, pits, feathers or other defects which detract from their appearance or impair their serviceability, proper functioning, or intended use.

Unit. Each (EA). One carver, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or or the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the metals, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one carver. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-2. For tests, the acceptance number shall be zero (0) for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be as specified herein. The carvers shall be examined for defects including, but not limited to, those listed below:

Defects

Style and design of instrument not as specified.
Finish of operative end and shank not as specified.
Finish of handle not as specified.
Dimensions of carver not as specified on drawing.
Edges and corners of handle not free of burrs and sharpness.
Plating of handle (brass handle only) not uniform in appearance.
Plating of handle (brass handle only) not free of blisters, visible porosity, or other defects.
Instrument not free from dirt, oil, polishing compound, or foreign matter.
Identification marking not present, not properly located, illegible, or incomplete.

Examination under magnification. The instruments shall be examined under four power (4x) magnification for the following defects:

Defects

Carving edges not properly sharpened.
Operative end not free from coarse grind marks, cracks, burrs, pits, feathers and other imperfections.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-282-7732	CARVER, DENTAL, Binangle, Cleoid-Discoid, No. UWD-5, Double-ended

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

CUSTODIANS:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No.: 6520-2292
Location: NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in letting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any *remarks* not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54211	CARVER, DENTAL, (Binangle, Cleoid-Discoid, No. UWD-5)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.