COMMERCIAL ITEM DESCRIPTION

SUPPORT, CERVICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers one type and four sizes of a two-piece, plastic and foam, radiolucent, cervical extrication collar, suitable for immobilization and maintenance of the cervical spine in a neutral position in order to prevent spinal cord injury during patient extrication.

Salient characteristics:

The support provided shall be in accordance with the size specified below:

Size	Description
1	Pediatric size, two-piece, cervical extrication support, fits neck circumference 8-16 inches, height 2-1/2 inches.
2	Small size, two-piece, cervical extrication support, fits neck circumference 10-20 inches, height 4 inches.
3	Medium size, two-piece, cervical extrication support, fits neck circumference 12-22 inches, height 4-3/4 inches.
4	Large size, two-piece, cervical extrication support, fits neck circumference 14-24 inches, height 5-1/2 inches.

General. The cervical support collar shall be two-pieces made of lightweight, flexible, 1/8 inch thick low density polyethylene, padded on the interior with 1/4 inch thick charcoal-colored, closed-cell foam padding to facilitate cleaning and promote reuse. The foam padding shall extend

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

past the borders of the polyetnylene on all parts of the collar. The foam shall be held to the rigid polyethylene by means of 5/8 inch diameter nylon rivets. The two-pieces of the cervical collar shall be adjustably coupled together by means of woven nylon hook and pile closures, located on both sides of the neck.

The anterior portion of the collar shall consist of two polyethylene panels. This anterior piece shall have a diamond shaped tracheal hole, a minimum of 2-1/2 inches in length and 1-1/2 inches in height which may be used for access to monitor carotid pulse or for airway management. Each anterior piece shall have a semi-circular, polyethylene chin piece designed to fold down flat for easy, compact storage. A strap of woven nylon pile shall extend from the side of the anterior panel and shall be used to hold the front piece in position while the posterior piece is applied. A strip of woven nylon hook shall extend across the length of the front panel.

The posterior portion of the collar shall be composed of two preformed polyethylene pieces riveted together in a minimum of three places. The smaller piece of the posterior panel shall be fully padded. The larger piece shall be lined along the bottom and sides with padding. Two pieces (straps) of woven nylon pile shall be riveted to the posterior panel on either side and shall be designed to attach the anterior piece to the posterior piece.

When coupled together, the support shall immobilize the cervical area by enclosing the neck, mandible, occiput, and the chin. The low density polyethylene construction shall permit the collar to conform to the patient. The collar shall provide rigid flexion, extension, axial and rotational immobilization. The cervical support shall maintain the spine in a neutral orientation during extrication and transport to prevent cervical spinal cord injury. The collar shall have clinical studies documenting capacity to immobilize.

The cervical support shall be packaged flat for ease of storage and may be reused. Each cervical support shall be furnished with an instruction sheet illustrating the sizing and application of the collar and describing the cleaning instructions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Workmanship. The cervical supports shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One individual package cervical support, as specified, constitutes one unit. Each individual package shall come with "Directions for Use".

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one cervical support as specified in this description.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one cervical support.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Fastener withstand test. The pile straps of the hook and pile fasteners used for application of the cervical support shall withstand a minimum pull of twenty-five (25) pounds, before detachment from the support.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	Item Identification
6515-01-263-7723	Support, Cervical, Extrication, Pediatric
6515-01-263-7724	Support, Cervical, Extrication, Small
6515-01-264-1376	Support, Cervical, Extrication, Medium
6515-01-263-7725	Support, Cervical, Extrication, Large

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-4763
Location:
CID/SUPPORT/CERVICAL/H15-37

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in ut ing comments and suggestions for improvements All users of military standardization documents are invited to provide sug ons. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54215	SUPPORT, CERVICAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify):

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.