COMMERCIAL ITEM DESCRIPTION

VIEWER, AGGLUTINATION, TEST TUBE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an agglutination viewer for use with hand held test tubes.

Salient Characteristics:

Shall be a magnified test tube agglutination viewer, for viewing blood typing, cross matching and other agglutination tests

Shall be designed for use with hand held test tubes

Shall consist of a shaded lamp with a 25 watt rated light bulb, an adjustable gooseneck, an adjustable magnifying mirror, an on-off electric switch, a lamp base and an electric power line cord

Shall be operable from 50/60 Hz, AC supply lines. The applicable line voltage shall be as indicated in the ordering data (see 'Ordering data' paragraph)

Shall be designed and constructed so that

Test tube(s) can be positioned between a top mounted lamp shade and a bottom mounted magnifying mirror

The content of the test tube(s) are illuminated from above.

The under surface contents of the test tube(s) are viewed below from the reflection in the magnifying mirror

Shall be supplied with a detachable 25 watt bulb that has a threaded base

Shall be constructed to include the following

A shaded lamp that is attached to the lamp base by a gooseneck

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

An on-off electric switch that is secured in a suitable cut-out in the lamp base

A magnifying mirror that is attached to the lamp base by a double universal ball joint. The mirror position shall be adjustable on the double universal ball joint and vary the magnification from 1-1/2X to 2X.

A minimum five (5) foot overall length line cord that terminates with a non-detachable three-blade male connector that is suitable for use in receptacles conforming to ANSI C73-11 or configuration No. 5-15P of NEMA Standard WD 6-88 The other end of the line cord shall be secured within the lamp base and shall be inserted into the lamp base through a suitable cutout that includes a grommet or bushing (for stain relief)

A lamp base that provides stability for the agglutination viewer when placed on a level and horizontal surface

Shall have all exposed metal parts coated with a suitable protective finish

Workmanship. The test tube agglutination viewer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One viewer as specified constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the test tube agglutination viewer in accordance with commercial practice to determine compliance with specification requirements. Sampling for the inspections and tests shall be conducted in accordance with the commercial practice. In addition, each test tube agglutination viewer shall be tested to assure proper electrical operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:31Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

C73-11 -Dimensions of Plugs and Receptacles-- General Purpose, 125 Volts, 15 Amperes, 2-Pole- 3-Wire Grounding Type

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway New York, NY 10018.)

NATIONAL ELECRTICAL MANUFACTURERS ASSOCIATION (NEMA)

WD 6-88 Wiring Devices - Dimensional Requirements (Revision and Redesignation of ANSI C73-73)

(Application for copies should be addressed to the National Electrical Manufacturers Association, 2101 L Street NW Suite 300, Washington DC 20037)

NOTE: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6640-01-165-6692	VIEWER, AGGLUTINATION, TEST TUBE, 115/230 Volt 50/60 Hz, AC

The following National Stock Number covers an incandescent lamp for use with a 115 Volt rated viewer

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6240-00-797-8945	LAMP INCANDESCENT 25 Watts 115 Volts

Ordering data. Applicable line voltage, intermediate/exterior package quantities and labeling and marking must be specified in the contract and or purchase order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1635

Location: ENABLE/VIEWERCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD invites you to use this form for submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edges (DO NOT STAPLE), and mailed. In block 8, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 8 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4066 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER
A-A-5422B

2. DOCUMENT TITLE
VIEWER, AGGLUTINATION, TEST TUBE

3a. NAME OF SUBMITTING ORGANIZATION

4. TYPE OF ORGANIZATION *(Mark one)*
- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*:

b. ADDRESS *(Street, City, State, ZIP Code)*

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional

b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional

c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 82 MAR

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.