A-A-54232
5 September 1990

COMMERCIAL ITEM DESCRIPTION

CATHETER, CARDIOVASCULAR (Arterial Embolectomy)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-42041(DM) which is cancelled.

This commercial item description covers a Fogarty cardiovascular catheter suitable for arterial embolectomy procedures

Salient characteristics:

General. The catheter shall be a cardiovascular catheter of the Fogarty design with a soft, atraumatic flexible tip and an inflatable concentric balloon

The catheters covered by this specification shall consist of the following syringe capacities and color codes:

Syringe Capacity		Color Code
1 cc	1	Green
3 cc	2	Red
Not Available(N/A)	3	White
	4	Blue
	5	Yellow

The balloon shall be securely affixed by epoxied sutures As an alternate, a suture tie sleeve for suture tying support may be used The proximal end shall have a gate valve and a female Luer lock hub in accordance with ANSI\HIMA MD70 1. The catheter shall be color coded to indicate catheter size The catheter may be supplied with a disposable lcc syringe except for the 6 and 7 Fr sizes The 3 Fr catheter may be supplied with a stylet if one is needed for stiffness The catheter shall be furnished with a stylet which extends throughout the length of the catheter Catheter shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 22298 Stainless Steel sleeves are required for catheters having French sizes of 4, 6 and 7, and optional for French sizes 3 and 5 Stylets are required for French sizes 4, 5, 6 and 7; and optional for French size 3

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Material The catheter shall be fabricated of polyvinylchloride which meets the requirements of the USP for Class IV plastics. The stylet and suture tie sleeve shall be fabricated of Type 304 stainless steel in accordance with ASTM Designation F899 The inflatable balloon shall be fabricated from latex rubber

Sterility, pyrogencity and toxicity The catheter shall be non-pyrogenic and nontoxic in accordance with the USP The catheter shall be sterile in accordance with USP or AAMI guidelines at a Sterility Assurance Level of 0 000001.

Radiopacity Except for the insertion depth markings and the latex balloon, the catheter shall be radiopaque in accordance with commercial practice

Saline resistance The catheter and balloon shall be tested for intravenous fluid compatibility by submersion in 5% USP saline solution at 37 +/- °C for 72 hours Afterward, the balloon shall be fully inflated and deflated once every 20 minutes for 8 hours

Shelf life. The shelf life date shall be printed on the package label and at least 24 months shelf life shall remain at the time of delivery to the Government.

Markings. The catheter shall be marked with product model number and balloon inflation capacity The catheter shall be legibly marked every 10 centimeters to indicate depth of penetration

Workmanship The catheter shall be free from defects which affect life, utility or appearance.

PRODUCT QUALITY LEVEL

Visual and dimensional examination Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Each (EA) Each catheter packaged in a rigid plastic tray or tube, as specified, constitutes one unit The balloon shall be provided with a protector The tray shall be packaged in a sealed, peel-open type container capable of maintaining sterility of the content unless the container is damaged or opened The container shall be constructed of tyvek having clear plastic backing

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The catheter shall be conducted to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one catheter. Sampling shall be inspection level S-2 with an AQL of 1.0

Contract certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers and their respective Item Identifications, Syringe Capacities, and Color Codes are covered by this document

NSN	ITEM IDENTIFICATION	SYRINGE CAPACITY	COLOR CODE
6515-01-115-1726	Catheter, Cardiovascular, Fogarty, Arterial Embolectomy, Plastic, 3 Fr, 80 cm	N/A	1
6515-01-139-9050	Catheter, Cardiovascular, Fogarty, Arterial Embolectomy, Plastic, 4 Fr, 80 cm	1 cc	2
6515-01-115-4710	Catheter, Cardiovascular, Fogarty, Arterial Embolectomy, Plastic, 5 Fr, 80 cm	N/A	3
6515-01-177-4924	Catheter, Cardiovascular, Fogarty, Arterial Embolectomy, Plastic, 6 Fr, 80 cm	3 cc	4
6515-01-174-9896	Catheter, Cardiovascular, Fogarty, Arterial Embolectomy, Plastic, 7 Fr, 80 cm	3 cc	5

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent.

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MFQAS

Project No 6515-4328

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54232	CATHETER, CARDIOVASCULAR (Arterial Embolectomy)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.