A-A-54234
30 August 1990

COMMERCIAL ITEM DESCRIPTION

CURETTES, PERIODONTAL, DOUBLE-ENDED

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-C-42000 and Interim Amendment 1 (DM) which are cancelled.

This commercial item description covers double-ended periodontal curettes suitable for use in subgingival curettage and the scaling of tooth and root surfaces..

Salient Characteristics:

Classification. The double-ended curettes shall be of the following styles:

Style 1 - Crane-Kaplan, No. CK-6.
Style 2 - Prichard, No. PR1 and PR2.
Style 3 - Sugarman, No. 5S and 6S.
Style 4 - Younger-Good, No. 7 and 8.
Style 5 - Kirkland, No. 4K and 5K.

Bid sample. Twelve bid samples which are representative of the curettes to be supplied shall be furnished complete with protocols and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center for evaluation by the Military Medical Services. When tested by the periodontists under conditions of normal usage, the curettes shall be suitable for use as follows:

Crane-Kaplan No. CK-6 Curette. The Crane-Kaplan, No. CK-6 curette, (Style 1), shall be suitable for use in removal of granulomatous tissue during periodontal surgical procedures. The curette shall be well balanced and blades finished to produce sharp cutting edges which shall be effective for tooth scaling and subgingival curettage of tooth and root surfaces with a mucoperiosteal flap approach.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Prichard No. PR1 and PR2 Curette. The Prichard No. PR1 and PR2 curette, (Style 2), shall be suitable for use in removal of granulomatous tissue during periodontal surgical procedures. The curette shall be well balanced and blades finished to produce sharp cutting edges which shall be effective for debridement with a mucoperiosteal flap approach. The curette shall be suitable when used on the root surface and to finish an ostectomy by removing thin bone on the tooth root after reduction of thickness with a bur.

Sugarman No. 5S and 6S Curette. The Sugarman No. 5S and 6S curette, (Style 3), shall be suitable for use in periodontal surgical procedures. The curettes shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade which shall effectively accomplish subgingival curettage. The curette shall also be suitable for osseous procedures requiring the removal of bone for autogenous implants.

Younger-Good No. 7 and 8 Curette. The Younger-Good No. 7 and 8 curette, (Style 4), shall be suitable for use in periodontal surgical procedures. The curettes shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade which shall effectively accomplish subgingival curettage and scaling of tooth and root surfaces.

Kirkland No. 4K and 5K Curette. The Kirkland No. 4K and 5K curette, (Style 5), shall be suitable for use in periodontal surgical and therapeutic procedures. The curette shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade. The curette shall be effective in removal of deposits and smoothing the tooth surfaces following tissue removal in the gingivectomy technique.

Material: Operative ends. The operative ends shall be fabricated of type 440A, 440B or 440C stainless steel. Stainless steel shall be in accordance with ASTM Designation F899 for composition when tested as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Material: Handle. Handle A and D shall be chromium over nickel plated brass and shall be hollow so as to provide lightness and thereby facilitate tactile sensibility. Brass shall conform to class 8, copper alloy number 260 or 330, as specified in ASTM Designation B135 for chemical composition only. Handle B shall be fabricated of type 416 or 410 stainless steel in accordance with ASTM Designation F899 or type 430F stainless steel in accordance with ASTM Designation A473 for composition. Handle C shall be fabricated of type 303 stainless steel as specified in ASTM Designation F899 for composition. Composition shall be tested as specified herein.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A, C54 to C58 for type 440B, and C56 to C60 for type 440C. Handle and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended. Hardness shall be determine in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Design and dimension. The curettes shall be double-ended and furnished in the applicable style and number specified. The curettes shall be suitable for use as previously specified. The curettes shall be furnished with handle and shank design of either Handle A, B, C or D as specified. The style, design and dimensions of the curettes shall conform to applicable Defense Personnel Support Center (DPSC) Drawings as shown below for the applicable styles.

The curette shall be packaged in a close fitting container and packaged to prevent damage to the operative end.

Style and Drawing Number

Style	DPSC Drawing No.
1-Crane-Kaplan, No. CK-6	24413
2-Prichard, No. PR1 and PR2	24413
3-Sugarman, No. 5S and 6S	24430
4-Younger-Good, No. 7 and 8	24440
5-Kirkland, No. 4K and 5K	24533

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Construction. The curettes shall be of three piece construction consisting of a handle and an operative end at each end. Handles shall be of solid or hollow construction as specified on the applicable drawing. The operative ends shall be securely attached to the handle. The interface of the shank and handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified herein.

Finish. The operative ends, including shanks, shall have a smooth mirror or smooth bright finish, except for the surfaces that form the cutting edges which shall have a finely honed finish. The operative ends (entire blade) shall have a sharp cutting edge and shall be free of defects such as coarse grind marks, cracks, gouges, burrs, feathers or other imperfections which affect curettaging performance, when viewed under four power (4X) magnification. Honed portions of the operative ends shall be blended at the shank junction to form a smooth contour. Handles shall have a mirror or bright finish, except for finger grip portions, which shall be serrated or knurled. Finger grip portions of handles shall be free from barbs and shall provide adequate finger grip during use.

Corrosion resistance. The copper sulfate test requirements of ASTM Designation F1089 shall apply to the operative ends. The boil test requirements of ASTM Designation F1089 shall apply to the entire curette (operative ends and handle).

Identification marking. Each curette shall be permanently and legibly marked with the applicable style name and instrument numbers. Each curette also shall be marked with the name or registered trade mark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless" shall be marked on the handle to identify the operative ends as stainless. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer. Curettes of foreign manufacture shall be marked with the country of origin. The manufacturer shall also provide a marking to trace the curettes to the contract and lot from which the instrument originated. All markings shall be on the smooth portion of the handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Workmanship. The operative ends shall be free of deep pits, discoloration and rounded edges. The curette shall be finished as specified and shall be free of all defects which detract from its appearance or impair its intended use.

Unit. Each (EA). One curette, of the style specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one curette, as specified in this description.

Examination. Examination shall be as specified herein. Examination shall include, but shall not be limited to the following:

Defects
Style and design not as specified for the applicable style and number curette.
Instrument not constructed as specified.
Operative end finish not as specified.
Honed portion of operative ends not blended at junction of shank to form a smooth contour.
Instrument not permanently and legibly marked with name or trademark of the contractor.
Instrument identification not present, properly located, or not complete.
Curette not free of dirt, oil, compound or foreign matter.
Handle serrations or knurling missing, not uniform or not complete.
Interface of shank and handle not smoothly blended with no crevices evident.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The inspection levels and acceptable quality levels (AQLs) shall be as indicated below. The unit of product for test purposes shall be one curette.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Test Criteria

Component	Characteristic	Inspection Level	AQL
Entire instrument	Chemical analysis	1/	2/
Operative ends	Hardness	S-2	1.5
Operative ends	Dimensions	S-2	2.5
Handle, shank, overall instrument	Dimensions	S-4	4.0
Operative ends and handle	Pull test	S-2	1.0
Cutting edges	Magnification Inspection	I	2/
Operative ends	Copper sulfate test	S-2	1.0
Entire instrument	Boil test	S-2	1.0

1/ A sufficient amount of each lot of metal used in the construction of the curettes may be selected for chemical analysis in lieu of sample curettes, provided lot identity has been maintained and can be demonstrated to the Government.

2/ Acceptance number shall be zero for all sample sizes.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Chemical analysis. In the absence of certificates of quality, chemical analysis of the individual components of the curette shall be determined in accordance with method 111 of FED-STD-151.

Dimensions. Dimensions shall be measured using a shadowgraph, measuring microscope, or other suitable device capable of measuring these dimensions with the accuracy specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end of the double-ended instrument.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirement.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-046-7333	CURETTE, PERIODONTAL, Crane-Kaplan, No. CK-6.
6520-01-046-7336	CURETTE, PERIODONTAL, Prichard, No. PR1 and PR2.
6520-01-093-2288	CURETTE, PERIODONTAL, Sugarman, No. 5S and 6S.
6520-01-093-2289	CURETTE, PERIODONTAL, Younger-Good, No 7 and 8.
6520-01-221-9077	CURETTE, PERIODONTAL, No. 4K and 5K.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC

PROJECT NO. 6520-2363
LOCATION: Word/PB/CDCUPDEN

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54234	CURETTES, PERIODONTAL, DOUBLE-ENDED

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.