METRIC

A-A-54244
3 September 1990

COMMERCIAL ITEM DESCRIPTION

FILES, PULP CANAL, DENTAL, (Hedstrom)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-F-0036782A (DM) which is cancelled.

This commercial item description covers standardized pulp canal files of hedstrom design for use in endodontic procedures.

Salient characteristics.

Except as noted below, the Hedstrom files shall be in accordance with ANSI (American National Standards Institute)/ADA (American Dental Association) Specification No. 58. In the case of a conflict this document shall govern.

Classification. Class I files shall have a 21 millimeter (mm) operative end length and shall be available in sizes 15, 20, 25, 30, 35, 40, 45, 50, 55, 60, 70, 80, 90 and 100. Class II files shall have a 25 mm operative end length and shall be available in sizes 25, 30, 35, 40, 45, 50, 55, 60, 70, 80, 90, and 100.

Material. The operative end of the file shall be fabricated from stainless steel type 302 or 304 in accordance with ASTM F899. The handle shall be fabricated from rigid plastic.

Style and design. Shall be a short-handled Hedstrom type pulp canal file designed as a cutting instrument for gaining access to, cleaning, enlarging and smoothing the walls of a root canal during endodontic procedures. The operative end of the file shall consist of a conical distal tip, a uniformly tapered spiral cutting portion and a non-tapered shank. The spirals forming the cutting edges shall be right handed and shall extend to the tip. The handle shall be designed to provide a secure finger grip. The handle shall be longitudinally serrated or shall have cross cut serrations. The entire handle shall be serrated or as an alternate, only the finger-gripping portion shall be serrated.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

The handle shall be in accordance with Figure 1. The files shall be packaged in a rigid container designed to prevent damage during shipment or storage.

Tensile strength. Sizes 15 through 80 files shall have a minimum tensile strength of 1860 M pascal. Sizes 90 and 100 files shall have a minimum tensile strength of 1634 M pascal. Tensile strength test shall be performed on a 20cm sample of steel stock used to produce the files.

Finish. The spiral section of the operative end of the file shall have a bright and smooth finish. The instrument shall be uniform and free from burrs, scales, rough areas and grind marks.

Identification marking. The handle shall be color coded and numbered with the respective size. Size and color shall be in accordance with ANSI/ADA specification No. 58.

Workmanship. The file shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing six files of pulp canal, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The unit of product for examination purposes shall be six files. Sampling for examination shall be conducted at inspection level II, with an acceptable quality level (AQL) of 2.5 except for dimensions which shall be conducted at an inspection level of S-1 with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall include but shall not be limited to the following:

Defects

Class and size not as specified.
Style and design not as specified.
Handle not properly marked with size or color coded.
Operative end finish not as specified.
Cutting edges not sharp.
Number of spirals not as specified.
Operative end bent, deformed or damaged.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. Unit of product for test purposes shall be six files. Sampling shall be conducted at an inspection level of S-1 with an acceptable quality level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.
Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquistion Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

A-A-

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
	FILE, PULP CANAL, DENTAL, 21 mm END LENGTH CLASS I
6520-00-935-4145	Hedstrom, Standardized, Size 15, 6s
6520-00-935-4146	Hedstrom, Standardized, Size 20, 6s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification
	FILE, PULP CANAL, DENTAL, 21 mm END LENGTH CLASS I (continued)
6520-00-935-4147	Hedstrom, Standardized, Size 25, 6s
6520-00-935-4148	Hedstrom, Standardized, Size 30, 6s
6520-00-935-4149	Hedstrom, Standardized, Size 35, 6s
6520-00-935-4150	Hedstrom, Standardized, Size 40, 6s
6520-00-935-4151	Hedstrom, Standardized, Size 45, 6s
6520-00-935-4152	Hedstrom, Standardized, Size 50, 6s
6520-00-935-4153	Hedstrom, Standardized, Size 55, 6s
6520-00-935-4154	Hedstrom, Standardized, Size 60, 6s
6520-00-935-4155	Hedstrom, Standardized, Size 70, 6s
6520-00-935-4156	Hedstrom, Standardized, Size 80, 6s
6520-00-935-4157	Hedstrom, Standardized, Size 90, 6s
6520-00-935-4158	Hedstrom, Standardized, Size 100, 6s
6520-00-935-4162	Hedstrom, Standardized, Size 25, 6s
6520-00-935-4163	Hedstrom, Standardized, Size 30, 6s
6520-00-935-4164	Hedstrom, Standardized, Size 35, 6s
6520-00-935-4165	Hedstrom, Standardized, Size 40, 6s
6520-00-935-4166	Hedstrom, Standardized, Size 45, 6s
6520-00-935-4167	Hedstrom, Standardized, Size 50, 6s
6520-00-935-4168	Hedstrom, Standardized, Size 55, 6s
6520-00-935-4169	Hedstrom, Standardized, Size 60, 6s
6520-00-935-4170	Hedstrom, Standardized, Size 70, 6s
6520-00-935-4171	Hedstrom, Standardized, Size 80, 6s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification
	FILE, PULP CANAL, DENTAL, 25 mm END LENGTH CLASS II (continued)
6520-00-935-4172	Hedstrom, Standardized, Size 90, 6s
6520-00-935-4173	Hedstrom, Standardized, Size 100, 6s

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM


CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2127


LOCATION: NCR Disk No. 7

05458

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.



HANDLE
(PLASTIC)



ALTERNATE HANDLE

NOTES: Handles shall be in strict accordance with this figure for dimensional requirements. The style and design of the handles shown above are typical and are not intended to restrict or preclude the purchase of files which are otherwise in accordance with the requirements of this commercial item description.

FIGURE 1: Handle, Files

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54244	FILES, PULP CANAL, DENTAL (Hedstrom)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.